

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/4/2013
Serial Number: 90-000210
AssetTag: 775-22001

Facility

Colorado Department of Health
Ship To:
Colorado Department of Health
8100 Lowrey Blvd
Denver, CO 80230

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99920 Pass ☒
Calibration Adjustment Pass ☒
Maintenance Log Entry ☒

Performed By: 467497 Date: 4/4/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1011 g/210L
☒ Linearity Test ☒ Interferant Detected
☒ Stability Test Average: 0.1008 g/210L
☒ Standard Deviation (<0.003) 0.0009
☒ Relative Standard Deviation (<3%) 0.0084
☒ EBAT Breath Test ☒ Final EBAT Test
☒ Improper Sample Test ☒ Maintenance Log Entry
☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com
☒ Card Reader ☒ Signature Pad
☒ IP Settings ☒ Modem
☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0210130001I

Performed By: 467497 Date: 4/5/2013

I-9000 Equipment Inventory

- ☐ Simulator Power Supplies (2)
- ☐ Keyboard
- ☐ Signature Pad with USB Adapter
- ☐ Simulator Communication Cable
- ☒ Printer Communication Cable
- ☐ Trainer Access Card
- ☐ Power Supply Cables (3)
- ☐ Ethernet Cable
- ☐ Standard Simulator Solution

Lot Number:

Performed By: 467497
Date: 4/5/2013

Documents

- ☒ Certificate
- ☐ Exception Message Guide
- ☐ Operator Guide
- ☐ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

- ☐ Printer

Reviewed By: 079490
Date: 4/16/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/5/2013
Serial Number: 90-000211
AssetTag: 775-22002

Facility

Larimer County Sheriff's Office
Ship To:
Larimer County Sheriff's Office
2405 Midpoint Drive
Fort Collins, CO 80525

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99936 Pass ☒
Calibration Adjustment Pass ☒
Maintenance Log Entry ☒

Performed By: 467497 Date: 4/5/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1009 g/210L
☒ Linearity Test ☒ Interferant Detected
☒ Stability Test Average: 0.1005 g/210L
☒ Standard Deviation (<0.003) 0.0008
☒ Relative Standard Deviation (<3%) 0.0082
☒ EBAT Breath Test ☒ Final EBAT Test
☒ Improper Sample Test ☒ Maintenance Log Entry
☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com
☒ Card Reader ☒ Signature Pad
☒ IP Settings ☒ Modem
☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0211130001I

Performed By: 467497 Date: 4/5/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)
☒ Keyboard
☒ Signature Pad with USB Adapter
☒ Simulator Communication Cable
☒ Printer Communication Cable
☒ Trainer Access Card
☒ Power Supply Cables (3)
☒ Ethernet Cable
☒ Standard Simulator Solution

Lot Number:

Performed By: 467497
Date: 4/5/2013

Documents

☒ Certificate
☒ Exception Message Guide
☒ Operator Guide
☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1526 MP1532

☒ Printer

Reviewed By: 079490
Date: 4/16/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/24/2013

Serial Number: 90-000212

AssetTag: 775-22003

Facility

Colorado Springs Police Department

Sand Creek

Ship To:

Colorado Springs Police Department

4125 Center Park Drive

Colorado Springs, CO 80916

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99942 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/24/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1007 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1004 g/210L

☒ Standard Deviation (<0.003) 0.0008

☒ Relative Standard Deviation (<3%) 0.0082

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0212130001I

Performed By: 732516 Date: 4/25/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/25/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1596

MP1643

☒ Printer

Reviewed By: 467497

Date: 4/26/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/17/2013

Serial Number: 90-000213

AssetTag: 775-22004

Facility

Loveland Police Department

Ship To:

Loveland Police Department

810 East 10th Street

Loveland, CO 80537

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99962 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/17/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0998 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0995 g/210L

☒ Standard Deviation (<0.003) .0007

☒ Relative Standard Deviation (<3%) .0069

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0213130001I

Performed By: 467497 Date: 4/18/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/18/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1587

MP1521

☒ Printer

Reviewed By: 079490

Date: 4/18/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/17/2013

Serial Number: 90-000214

AssetTag: 775-22005

Facility

Eagle County Sheriff's Office

Jail

Ship To:

Eagle County Sheriff's Office

885 East Chambers Avenue

Eagle, CO 81631

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99923 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/17/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0985 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0987 g/210L

☒ Standard Deviation (<0.003) 0.0005

☒ Relative Standard Deviation (<3%) 0.0048

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0214130001I

Performed By: 467497 Date: 4/18/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/18/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1390

MP1598

☒ Printer

Reviewed By: 079490

Date: 4/18/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/16/2013

Serial Number: 90-000216

AssetTag: 775-22007

Facility

Colorado Department of Health

Ship To:

Colorado Department of Health

8100 Lowrey Blvd

Denver, CO 80230

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99966 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 079490 Date: 4/16/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0997 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0997 g/210L

☒ Standard Deviation (<0.003) 0.0007

☒ Relative Standard Deviation (<3%) 0.0075

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0216130001I

Performed By: 079490 Date: 4/17/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 079490

Date: 4/17/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

☒ Printer

Reviewed By: 467497

Date: 4/17/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/24/2013

Serial Number: 90-000218

AssetTag: 775-22009

Facility

Pitkin County Sheriff's Office

Ship To:

Pitkin County Sheriff's Office

506 E. Main Street, Suite 101

Aspen, CO 81611-2923

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99999 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/25/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0984 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0977 g/210L

☒ Standard Deviation (<0.003) 0.0008

☒ Relative Standard Deviation (<3%) 0.0083

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0218130001I

Performed By: 732516 Date: 4/25/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/25/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1642

MP1606

☒ Printer

Reviewed By: 467497

Date: 4/26/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/24/2013

Serial Number: 90-000219

AssetTag: 775-22010

Facility

Rio Grande County Sheriff's Office

Ship To:

Rio Grande County Sheriff's Office

640 Cherry Street

Del Norte, CO 81132

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) .99960 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/24/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0995 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0990 g/210L

☒ Standard Deviation (<0.003) .0008

☒ Relative Standard Deviation (<3%) .0077

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0219130001i

Performed By: 732516 Date: 4/24/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/24/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1413

MP1605

☒ Printer

Reviewed By: 079490

Date: 4/24/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/11/2013

Serial Number: 90-000220

AssetTag: 775-22011

Facility

Fort Lupton Police Department

Ship To:

Fort Lupton Police Department

130 South McKinley Avenue

Fort Lupton, CO 80621

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) .99996 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/11/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.103 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1020 g/210L

☒ Standard Deviation (<0.003) .0007

☒ Relative Standard Deviation (<3%) .0071

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0220130001I

Performed By: 467497 Date: 4/11/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/11/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1568

MP1364

☒ Printer

Reviewed By: 079490

Date: 4/16/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/17/2013

Serial Number: 90-000222

AssetTag: 775-22013

Facility

Las Animas County Sheriff's Office

Ship To:

Las Animas County Sheriff's Office

2309 East Main Street

Trinidad, CO 81082

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99899 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/17/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0984 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0986 g/210L

☒ Standard Deviation (<0.003) 0.0006

☒ Relative Standard Deviation (<3%) 0.0061

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0222130001I

Performed By: 467497 Date: 4/18/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/18/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1321

MP1358

☒ Printer

Reviewed By: 079490

Date: 4/18/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/11/2013

Serial Number: 90-000224

AssetTag: 775-22015

Facility

Fountain Police Department

Ship To:

Fountain Police Department

222 North Santa Fe Avenue

Fountain, CO 80817

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99897 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/11/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1011 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0988 g/210L

☒ Standard Deviation (<0.003) 0.0009

☒ Relative Standard Deviation (<3%) .0091

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0224130001I

Performed By: 467497 Date: 4/11/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/11/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1507

MP1573

☒ Printer

Reviewed By: 079490

Date: 4/16/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/11/2013

Serial Number: 90-000225

AssetTag: 775-22016

Facility

Greenwood Village Police Department

Ship To:

Greenwood Village Police Department

6060 South Quebec Street

Greenwood Village, CO 80111-4591

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99957 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/11/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1014 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1010 g/210L

☒ Standard Deviation (<0.003) 0.0007

☒ Relative Standard Deviation (<3%) 0.0072

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0225130001I

Performed By: 467497 Date: 4/11/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/11/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1375

MP1386

☒ Printer

Reviewed By: 079490

Date: 4/16/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/24/2013

Serial Number: 90-000226

AssetTag: 775-22017

Facility

Aurora Police Department

1 Jail

Ship To:

Aurora Police Department
15001 East Alameda Drive
Aurora, CO 80012-1547

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99853 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/24/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1004 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1009 g/210L

☒ Standard Deviation (<0.003) 0.0007

☒ Relative Standard Deviation (<3%) 0.0067

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0226130001I

Performed By: 732516 Date: 4/24/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/24/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1359

MP1601

☒ Printer

Reviewed By: 079490

Date: 4/24/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/17/2013

Serial Number: 90-000227

AssetTag: 775-22018

Facility

Lake County Sheriff's Office

Ship To:

Lake County Sheriff's Office

500 Harrison Avenue

Leadville, CO 80461

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99930 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/17/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0991 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0988 g/210L

☒ Standard Deviation (<0.003) 0.0007

☒ Relative Standard Deviation (<3%) 0.0073

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0227130001I

Performed By: 467497 Date: 4/18/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/18/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

☒ Printer

Reviewed By: 079490

Date: 4/18/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/12/2013

Serial Number: 90-000230

AssetTag: 775-22021

Facility

Cortez Police Department

Ship To:

Cortez Police Department

608 North Park Street

Cortez, CO 81321

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99830 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/12/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1002 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1003 g/210L

☒ Standard Deviation (<0.003) .0007

☒ Relative Standard Deviation (<3%) .0073

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0230130001I

Performed By: 467497 Date: 4/12/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/12/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1425

MP1555

☒ Printer

Reviewed By: 079490

Date: 4/16/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/11/2013

Serial Number: 90-000231

AssetTag: 775-22022

Facility

Prowers County Sheriff's Office

Ship To:

Prowers County Sheriff's Office

103 East Oak Street

Lamar, CO 81052

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99963 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/11/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1029 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1022 g/210L

☒ Standard Deviation (<0.003) 0.0005

☒ Relative Standard Deviation (<3%) .0048

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0231130001I

Performed By: 467497 Date: 4/11/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/11/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1459

MP1533

☒ Printer

Reviewed By: 079490

Date: 4/16/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/17/2013

Serial Number: 90-000232

AssetTag: 775-22023

Facility

Pueblo County Sheriff's Office

Ship To:

Pueblo County Sheriff's Office

909 Court Street

Pueblo, CO 81003

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99936 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/17/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1010 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1006 g/210L

☒ Standard Deviation (<0.003) 0.0018

☒ Relative Standard Deviation (<3%) 0.0175

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0232130001I

Performed By: 467497 Date: 4/18/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/18/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1570

MP1508

☒ Printer

Reviewed By: 079490

Date: 4/18/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/6/2013
Serial Number: 90-000233
AssetTag: 775-22024

Facility

Otero County Sheriff's Office
Ship To:
Otero County Sheriff's Office
222 East 2nd Street
La Junta, CO 81050

MAC Address:
IP Address:
Gateway:
Subnet Mask:
Analog Phone:

Calibration Adjust
Flow Sensor (>0.998) 0.99948 Pass ☒
Calibration Adjustment Pass ☒
Maintenance Log Entry ☒ Performed By: 079490 Date: 4/16/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1001 g/210L
☒ Linearity Test ☒ Interferant Detected
☒ Stability Test Average: 0.1011 g/210L
☒ Standard Deviation (<0.003) 0.0008
☒ Relative Standard Deviation (<3%) 0.0075
☒ EBAT Breath Test ☒ Final EBAT Test
☒ Improper Sample Test ☒ Maintenance Log Entry
☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com
☒ Card Reader ☒ Signature Pad
☒ IP Settings ☒ Modem
☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #
0233130001I
Performed By: 079490 Date: 4/17/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)
☒ Keyboard
☒ Signature Pad with USB Adapter
☒ Simulator Communication Cable
☒ Printer Communication Cable
☒ Trainer Access Card
☒ Power Supply Cables (3)
☒ Ethernet Cable
☒ Standard Simulator Solution
Lot Number:

Performed By: 079490
Date: 4/17/2013

Documents

☒ Certificate
☒ Exception Message Guide
☒ Operator Guide
☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)
MP1680 MP1469
☒ Printer

Reviewed By: 467497
Date: 4/17/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/11/2013

Serial Number: 90-000234

AssetTag: 775-22025

Facility

Montezuma County Sheriff's Office

Ship To:

Montezuma County Sheriff's Office

730 East Driscoll

Cortez, CO 81321

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99932 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/11/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1013 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1015 g/210L

☒ Standard Deviation (<0.003) .0007

☒ Relative Standard Deviation (<3%) 0.0068

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0234130001I

Performed By: 467497 Date: 4/11/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/11/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1434

MP1612

☒ Printer

Reviewed By: 079490

Date: 4/16/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/19/2013

Serial Number: 90-000235

AssetTag: 775-22026

Facility

Adams County Sheriff's Office

CSP Strasburg

Ship To:

Adams County Sheriff's Office

2550 Strasburg Mile Rd.

Strasburg, CO 80136

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99931 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/19/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0989 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0985 g/210L

☒ Standard Deviation (<0.003) 0.0006

☒ Relative Standard Deviation (<3%) 0.0061

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0234130001I

Performed By: 467497 Date: 4/20/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/20/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

☒ Printer

Reviewed By: 079490

Date: 4/22/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/9/2013
Serial Number: 90-000237
AssetTag: 775-22028

Facility

Denver Police Department
Ship To:
Denver Police Department
3381 Park Avenue West
Denver, CO 80216

Bat Van

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99999 Pass ☒
Calibration Adjustment Pass ☒
Maintenance Log Entry ☒

Performed By: 467497 Date: 4/9/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0996 g/210L
☒ Linearity Test ☒ Interferant Detected
☒ Stability Test Average: 0.0995 g/210L
☒ Standard Deviation (<0.003) .0009
☒ Relative Standard Deviation (<3%) .0089
☒ EBAT Breath Test ☒ Final EBAT Test
☒ Improper Sample Test ☒ Maintenance Log Entry
☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com
☒ Card Reader ☒ Signature Pad
☒ IP Settings ☒ Modem
☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0237130001I

Performed By: 467497 Date: 4/9/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)
☒ Keyboard
☒ Signature Pad with USB Adapter
☒ Simulator Communication Cable
☒ Printer Communication Cable
☒ Trainer Access Card
☒ Power Supply Cables (3)
☒ Ethernet Cable
☒ Standard Simulator Solution

Lot Number:

Performed By: 467497
Date: 4/9/2013

Documents

☒ Certificate
☒ Exception Message Guide
☒ Operator Guide
☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1334 MP1430

☒ Printer

Reviewed By: 079490
Date: 4/20/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/11/2013

Serial Number: 90-000238

AssetTag: 775-22029

Facility

Colorado Springs Police Department

BAT Van

Ship To:

Colorado Springs Police Department

705 South Nevada Avenue

Colorado Springs, CO 80903

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99829 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/11/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1030 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1015 g/210L

☒ Standard Deviation (<0.003) .0005

☒ Relative Standard Deviation (<3%) .0051

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0238130001I

Performed By: 467497 Date: 4/11/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/11/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1669

MP1630

☒ Printer

Reviewed By: 079490

Date: 4/20/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/11/2013

Serial Number: 90-000239

AssetTag: 775-22030

Facility

Mesa County Sheriff's Office

Bat Van

Ship To:

Mesa County Sheriff's Office

215 Rice Street

Grand Junction, CO 81501

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99828 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/11/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1018 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1024 g/210L

☒ Standard Deviation (<0.003) .0006

☒ Relative Standard Deviation (<3%) .0058

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0239130001I

Performed By: 467497 Date: 4/11/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/11/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1629

MP1481

☒ Printer

Reviewed By: 079490

Date: 4/20/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/11/2013

Serial Number: 90-000240

AssetTag: 775-22031

Facility

Weld County Sheriff's Office

Bat Van

Ship To:

Weld County Sheriff's Office

1950 O Street

Greeley, CO 80631

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99936 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/11/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1029 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1017 g/210L

☒ Standard Deviation (<0.003) .0007

☒ Relative Standard Deviation (<3%) .0065

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0240130001I

Performed By: 467497 Date: 4/11/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/11/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1610

MP1485

☒ Printer

Reviewed By: 079490

Date: 4/20/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/17/2013

Serial Number: 90-000241

AssetTag: 775-22032

Facility

San Miguel County Sheriff's Office

Ship To:

San Miguel County Sheriff's Office

684 CR 63L

Telluride, CO 81435

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99909 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/17/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0997 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0991 g/210L

☒ Standard Deviation (<0.003) 0.0007

☒ Relative Standard Deviation (<3%) 0.0075

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0241130001I

Performed By: 467497 Date: 4/18/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/18/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1644

MP1565

☒ Printer

Reviewed By: 079490

Date: 4/18/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/24/2013

Serial Number: 90-000246

AssetTag: 775-22037

Facility

Durango Police Department

Ship To:

Durango Police Department

990 East Second Avenue

Durango, CO 81301

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99716 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/24/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.996 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.995 g/210L

☒ Standard Deviation (<0.003) 0.0008

☒ Relative Standard Deviation (<3%) 0.0083

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0246130001I

Performed By: 732516 Date: 4/24/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/24/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1640

MP1665

☒ Printer

Reviewed By: 467497

Date: 4/24/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/24/2013

Serial Number: 90-000248

AssetTag: 775-22039

Facility

Glendale Police Department

Ship To:

Glendale Police Department

950 South Birch Street

Glendale, CO 80222

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) .99905 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/24/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0999 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1007 g/210L

☒ Standard Deviation (<0.003) .0007

☒ Relative Standard Deviation (<3%) .0065

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0248130001I

Performed By: 732516 Date: 4/24/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/24/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1378

MP1600

☒ Printer

Reviewed By: 079490

Date: 4/24/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/24/2013

Serial Number: 90-000250

AssetTag: 775-22041

Facility

Calhan Police Department

Ship To:

Calhan Police Department

556 Colorado Avenue

Calhan, CO 80808

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99908 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/24/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0999 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1001 g/210L

☒ Standard Deviation (<0.003) 0.0006

☒ Relative Standard Deviation (<3%) 0.0060

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0250130001I

Performed By: 732516 Date: 4/24/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/24/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1569

MP1348

☒ Printer

Reviewed By: 079490

Date: 4/24/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/14/2013

Serial Number: 90-000251

AssetTag: 775-22042

Facility

Rangely Police Department

Ship To:

Rangely Police Department

209 East Main Street

Rangely, CO 81648

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99981 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/25/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0984 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0989 g/210L

☒ Standard Deviation (<0.003) .0007

☒ Relative Standard Deviation (<3%) .0068

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

025113001I

Performed By: 732516 Date: 4/25/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/25/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1412

MP1648

☒ Printer

Reviewed By: 467497

Date: 4/26/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/17/2013

Serial Number: 90-000253

AssetTag: 775-22044

Facility

Jackson County Sheriff's Office

Ship To:

Jackson County Sheriff's Office

396 La Fever

Walden, CO 80480

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99960 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/17/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0982 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0986 g/210L

☒ Standard Deviation (<0.003) 0.0006

☒ Relative Standard Deviation (<3%) 0.0061

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0253130001I

Performed By: 467497 Date: 4/18/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/18/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1624

MP1370

☒ Printer

Reviewed By: 079490

Date: 4/18/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/9/2013
Serial Number: 90-000254
AssetTag: 775-22045

Facility

Park County Sheriff's Office
Ship To:
Park County Sheriff's Office
59865 Highway 285
Bailey, CO 9880

Bailey

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99937 Pass ☒
Calibration Adjustment Pass ☒
Maintenance Log Entry ☒

Performed By: 467497 Date: 4/10/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0988 g/210L
☒ Linearity Test ☒ Interferant Detected
☒ Stability Test Average: 0.0986 g/210L
☒ Standard Deviation (<0.003) 0.0005
☒ Relative Standard Deviation (<3%) 0.0051
☒ EBAT Breath Test ☒ Final EBAT Test
☒ Improper Sample Test ☒ Maintenance Log Entry
☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com
☒ Card Reader ☒ Signature Pad
☒ IP Settings ☒ Modem
☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0254130001I

Performed By: 467497 Date: 4/10/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)
☒ Keyboard
☒ Signature Pad with USB Adapter
☒ Simulator Communication Cable
☒ Printer Communication Cable
☒ Trainer Access Card
☒ Power Supply Cables (3)
☒ Ethernet Cable
☒ Standard Simulator Solution

Lot Number:

Performed By: 467497
Date: 4/10/2013

Documents

☒ Certificate
☒ Exception Message Guide
☒ Operator Guide
☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

☒ Printer

Reviewed By: 079490
Date: 4/17/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/17/2013

Serial Number: 90-000255

AssetTag: 775-22046

Facility

Adams County Sheriff's Office

Ship To:

Adams County Sheriff's Office

4201 East 72nd Avenue

Commerce City, CO 80022

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99960 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/17/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0994 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.983 g/210L

☒ Standard Deviation (<0.003) 0.0006

☒ Relative Standard Deviation (<3%) 0.0065

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0255130001I

Performed By: 467497 Date: 4/17/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/17/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1441

MP1488

☒ Printer

Reviewed By: 079490

Date: 4/18/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/23/2013

Serial Number: 90-000256

AssetTag: 775-22047

Facility

Adams County Sheriff's Office

CSP Strasburg

Ship To:

Adams County Sheriff's Office

2550 Strasburg Mile Rd.

Strasburg, CO 80136

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99994 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 079490 Date: 4/23/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1001 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0993 g/210L

☒ Standard Deviation (<0.003) .0006

☒ Relative Standard Deviation (<3%) .0055

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0256130001I

Performed By: 079490 Date: 4/23/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 079490

Date: 4/23/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1460

MP1314

☒ Printer

Reviewed By: 467497

Date: 4/23/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/29/2013

Serial Number: 90-000257

AssetTag: 775-22048

Facility

460th SFS Buckley AFB

Ship To:

460th SFS Buckley AFB

18300 East Crested Butte, Stop 87, Bldg 706

Buckley Air Force Base, CO 80011

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) .99989 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/29/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1003 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1003 g/210L

☒ Standard Deviation (<0.003) 0.0006

☒ Relative Standard Deviation (<3%) 0.0057

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0257130001I

Performed By: 732516 Date: 4/29/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/29/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

☒ Printer

Reviewed By: 467497

Date: 4/29/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/26/2013

Serial Number: 90-000258

AssetTag: 775-22049

Facility

Arapahoe County Sheriff's Office

Ship To:

Arapahoe County Sheriff's Office

13101 East Broncos Parkway

Centennial, CO 80112

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99999 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/26/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0981 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0982 g/210L

☒ Standard Deviation (<0.003) .0008

☒ Relative Standard Deviation (<3%) .0085

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0258130001I

Performed By: 732516 Date: 4/26/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/26/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1336

MP1552

☒ Printer

Reviewed By: 079490

Date: 4/26/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/23/2013

Serial Number: 90-000259

AssetTag: 775-22050

Facility

Archuleta County Sheriff's Office

Ship To:

Archuleta County Sheriff's Office

551 Hot Springs Boulevard

Pagosa Springs, CO 81147

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99886 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/23/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0993 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0988 g/210L

☒ Standard Deviation (<0.003) .0004

☒ Relative Standard Deviation (<3%) .0045

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0259130001I

Performed By: 732516 Date: 4/23/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/23/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1423

MP1391

☒ Printer

Reviewed By: 079490

Date: 4/23/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/24/2013

Serial Number: 90-000260

AssetTag: 775-22051

Facility

Arvada Police Department

Ship To:

Arvada Police Department

5790 Garrison Street

Arvada, CO 80002

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99957 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/24/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1008 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1005 g/210L

☒ Standard Deviation (<0.003) 0.0006

☒ Relative Standard Deviation (<3%) 0.0060

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0260130001i

Performed By: 732516 Date: 4/25/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/25/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1333

MP1561

☒ Printer

Reviewed By: 467497

Date: 4/26/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/9/2013
Serial Number: 90-000261
AssetTag: 775-22052

Facility

Ault Police Department
Ship To:
Ault Police Department
203 1st Street
Ault, CO 80610

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99943 Pass ☒
Calibration Adjustment Pass ☒
Maintenance Log Entry ☒

Performed By: 467497 Date: 4/9/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0981 g/210L
☒ Linearity Test ☒ Interferant Detected
☒ Stability Test Average: 0.0993 g/210L
☒ Standard Deviation (<0.003) .0012
☒ Relative Standard Deviation (<3%) .0117
☒ EBAT Breath Test ☒ Final EBAT Test
☒ Improper Sample Test ☒ Maintenance Log Entry
☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com
☒ Card Reader ☒ Signature Pad
☒ IP Settings ☒ Modem
☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0261130001I

Performed By: 467497 Date: 4/9/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)
☒ Keyboard
☒ Signature Pad with USB Adapter
☒ Simulator Communication Cable
☒ Printer Communication Cable
☒ Trainer Access Card
☒ Power Supply Cables (3)
☒ Ethernet Cable
☒ Standard Simulator Solution

Lot Number:

Performed By: 467497
Date: 4/9/2013

Documents

☒ Certificate
☒ Exception Message Guide
☒ Operator Guide
☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1553 MP1419

☒ Printer

Reviewed By: 079490
Date: 4/17/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/10/2013

Serial Number: 90-000263

AssetTag: 775-22054

Facility

Avon Police Department

Ship To:

Avon Police Department

400 Bench Mark Road

Avon, CO 81620

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99899 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/10/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0969 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0977 g/210L

☒ Standard Deviation (<0.003) .0007

☒ Relative Standard Deviation (<3%) .0069

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0263130001I

Performed By: 467497 Date: 4/10/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/10/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1323

MP1452

☒ Printer

Reviewed By: 079490

Date: 4/20/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/10/2013

Serial Number: 90-000264

AssetTag: 775-22055

Facility

Baca County Sheriff's Office

Ship To:

Baca County Sheriff's Office

265 East 2nd Street

Springfield, CO 81073

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.999839 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/10/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0987 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0989 g/210L

☒ Standard Deviation (<0.003) .0004

☒ Relative Standard Deviation (<3%) .0040

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0264130001I

Performed By: 467497 Date: 4/10/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/10/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1396

MP1575

☒ Printer

Reviewed By: 079490

Date: 4/20/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/10/2013

Serial Number: 90-000265

AssetTag: 775-22056

Facility

Bent County Sheriff's Office

Ship To:

Bent County Sheriff's Office

11100 COUNTY ROAD GG.5

Las Animas, CO 81054

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99912 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/10/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0977 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0978 g/210L

☒ Standard Deviation (<0.003) .0010

☒ Relative Standard Deviation (<3%) .0103

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0265130001I

Performed By: 467497 Date: 4/10/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/10/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1422

MP1420

☒ Printer

Reviewed By: 079490

Date: 4/20/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/8/2013
Serial Number: 90-000266
AssetTag: 775-22057

Facility

Black Hawk Police Department
Ship To:
Black Hawk Police Department
221 Church Street
Black Hawk, CO 80422-0017

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99881 Pass ☒
Calibration Adjustment Pass ☒
Maintenance Log Entry ☒

Performed By: 467497 Date: 4/8/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1004 g/210L
☒ Linearity Test ☒ Interferant Detected
☒ Stability Test Average: 0.1000 g/210L
☒ Standard Deviation (<0.003) .0005
☒ Relative Standard Deviation (<3%) .0046
☒ EBAT Breath Test ☒ Final EBAT Test
☒ Improper Sample Test ☒ Maintenance Log Entry
☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com
☒ Card Reader ☒ Signature Pad
☒ IP Settings ☒ Modem
☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0266130001I

Performed By: 467497 Date: 4/9/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)
☒ Keyboard
☒ Signature Pad with USB Adapter
☒ Simulator Communication Cable
☒ Printer Communication Cable
☒ Trainer Access Card
☒ Power Supply Cables (3)
☒ Ethernet Cable
☒ Standard Simulator Solution

Lot Number:

Performed By: 467497
Date: 4/9/2013

Documents

☒ Certificate
☒ Exception Message Guide
☒ Operator Guide
☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1503 MP1468

☒ Printer

Reviewed By: 079490
Date: 4/17/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/5/2013
Serial Number: 90-000267
AssetTag: 775-22058

Facility

Boulder County Sheriff's Office
Ship To:
Boulder County Sheriff's Office
5600 Flat Iron Parkway
Boulder, CO 80301

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99990 Pass ☒
Calibration Adjustment Pass ☒
Maintenance Log Entry ☒

Performed By: 467497 Date: 4/5/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1004 g/210L
☒ Linearity Test ☒ Interferant Detected
☒ Stability Test Average: 0.1004 g/210L
☒ Standard Deviation (<0.003) .0007
☒ Relative Standard Deviation (<3%) .0068
☒ EBAT Breath Test ☒ Final EBAT Test
☒ Improper Sample Test ☒ Maintenance Log Entry
☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com
☒ Card Reader ☒ Signature Pad
☒ IP Settings ☒ Modem
☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0267130001I

Performed By: 467497 Date: 4/5/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)
☒ Keyboard
☒ Signature Pad with USB Adapter
☒ Simulator Communication Cable
☒ Printer Communication Cable
☒ Trainer Access Card
☒ Power Supply Cables (3)
☒ Ethernet Cable
☒ Standard Simulator Solution

Lot Number:

Performed By: 467497
Date: 4/5/2013

Documents

☒ Certificate
☒ Exception Message Guide
☒ Operator Guide
☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1403 MP1362

☒ Printer

Reviewed By: 079490
Date: 4/16/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/10/2013

Serial Number: 90-000268

AssetTag: 775-22059

Facility

Boulder Police Department

Ship To:

Boulder Police Department

1805 33rd Street

Boulder, CO 80302

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99861 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/10/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0993 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1012 g/210L

☒ Standard Deviation (<0.003) .0012

☒ Relative Standard Deviation (<3%) .0117

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0268130001I

Performed By: 467497 Date: 4/10/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/10/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1356

MP1684

☒ Printer

Reviewed By: 079490

Date: 4/20/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/12/2013

Serial Number: 90-000269

AssetTag: 775-22060

Facility

Broomfield Police Department

Jail

Ship To:

Broomfield Police Department

11600 Ridge Parkway

Broomfield, CO 80020

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99823 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/12/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1005 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1018 g/210L

☒ Standard Deviation (<0.003) .0008

☒ Relative Standard Deviation (<3%) .0075

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0269130001I

Performed By: 732516 Date: 4/12/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/12/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1517

MP1564

☒ Printer

Reviewed By: 079490

Date: 4/19/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/10/2013

Serial Number: 90-000271

AssetTag: 775-22062

Facility

Buena Vista Police Department

Ship To:

Buena Vista Police Department

123 Linderman Avenue

Buena Vista, CO 81211

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99962 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/10/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1003 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1012 g/210L

☒ Standard Deviation (<0.003) 0.0008

☒ Relative Standard Deviation (<3%) 0.0082

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0271130001I

Performed By: 467497 Date: 4/11/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/11/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1657

MP1463

☒ Printer

Reviewed By: 079490

Date: 4/16/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/19/2013

Serial Number: 90-000272

AssetTag: 775-22063

Facility

Carbondale Police Department
Ship To:
Carbondale Police Department
511 Colorado Avenue, Suite 911
Carbondale, CO 81623

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99935 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/19/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0991 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0987 g/210L

☒ Standard Deviation (<0.003) .0007

☒ Relative Standard Deviation (<3%) .0067

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0272130001I

Performed By: 732516 Date: 4/19/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/19/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1456

MP1634

☒ Printer

Reviewed By: 467497

Date: 4/22/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/22/2013

Serial Number: 90-000273

AssetTag: 775-22064

Facility

Castle Rock Police Department

Ship To:

Castle Rock Police Department

100 North Perry Street

Castle Rock, CO 80104-2486

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99850 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/22/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0991 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0985 g/210L

☒ Standard Deviation (<0.003) .0006

☒ Relative Standard Deviation (<3%) .0061

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0273130001I

Performed By: 732516 Date: 4/22/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/22/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1316

MP1409

☒ Printer

Reviewed By: 467497

Date: 4/22/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/17/2013

Serial Number: 90-000274

AssetTag: 775-22065

Facility

Center Police Department

Ship To:

Center Police Department

294 South Worth Street

Center, CO 81125

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99949 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/17/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0987 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0983 g/210L

☒ Standard Deviation (<0.003) .0004

☒ Relative Standard Deviation (<3%) .0045

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0274130001I

Performed By: 467497 Date: 4/17/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/17/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1591

MP1633

☒ Printer

Reviewed By: 079490

Date: 4/23/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/19/2013

Serial Number: 90-000275

AssetTag: 775-22066

Facility

Chaffee County Sheriff's Office

Ship To:

Chaffee County Sheriff's Office

132 Crestone

Salida, CO 81201

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99985 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/19/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0975 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0972 g/210L

☒ Standard Deviation (<0.003) .0006

☒ Relative Standard Deviation (<3%) .0060

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0275130001I

Performed By: 732516 Date: 4/19/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/19/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1635

MP1663

☒ Printer

Reviewed By: 079490

Date: 4/22/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/24/2013

Serial Number: 90-000276

AssetTag: 775-22067

Facility

Cherry Hills Police Department

Ship To:

Cherry Hills Police Department

2460 East Quincy Avenue

Englewood, CO 80113

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99962 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/24/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0982 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0971 g/210L

☒ Standard Deviation (<0.003) 0.0005

☒ Relative Standard Deviation (<3%) 0.0053

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0276130001I

Performed By: 732516 Date: 4/25/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/25/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1384

MP1509

☒ Printer

Reviewed By: 467497

Date: 4/26/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/23/2013

Serial Number: 90-000277

AssetTag: 775-22068

Facility

Cheyenne County Sheriff's Office

Ship To:

Cheyenne County Sheriff's Office

91 East 1st Street

Cheyenne Wells, CO 80810

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99968 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/23/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1004 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1005 g/210L

☒ Standard Deviation (<0.003) .0005

☒ Relative Standard Deviation (<3%) .0051

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0277130001I

Performed By: 732516 Date: 4/23/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/23/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1464

MP1466

☒ Printer

Reviewed By: 079490

Date: 4/23/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/23/2013

Serial Number: 90-000278

AssetTag: 775-22069

Facility

Clear Creek County Sheriff's Office

Ship To:

Clear Creek County Sheriff's Office

405 Argentine

GeorgeTown, CO 80444

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99987 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 079490 Date: 4/23/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0977 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0974 g/210L

☒ Standard Deviation (<0.003) .0006

☒ Relative Standard Deviation (<3%) .0061

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0278130001I

Performed By: 079490 Date: 4/23/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 079490

Date: 4/23/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1618

MP1686

☒ Printer

Reviewed By: 467497

Date: 4/23/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/9/2013
Serial Number: 90-000279
AssetTag: 775-22070

Facility

Colorado State Patrol
Ship To:
Colorado State Patrol
8200 North Highway 85
Commerce City, CO 80022

Adams County

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99949 Pass ☒
Calibration Adjustment Pass ☒
Maintenance Log Entry ☒

Performed By: 467497 Date: 4/9/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0977 g/210L
☒ Linearity Test ☒ Interferant Detected
☒ Stability Test Average: 0.0982 g/210L
☒ Standard Deviation (<0.003) .0007
☒ Relative Standard Deviation (<3%) .0071
☒ EBAT Breath Test ☒ Final EBAT Test
☒ Improper Sample Test ☒ Maintenance Log Entry
☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com
☒ Card Reader ☒ Signature Pad
☒ IP Settings ☒ Modem
☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0279130001I

Performed By: 467497 Date: 4/9/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)
☒ Keyboard
☒ Signature Pad with USB Adapter
☒ Simulator Communication Cable
☒ Printer Communication Cable
☒ Trainer Access Card
☒ Power Supply Cables (3)
☒ Ethernet Cable
☒ Standard Simulator Solution

Lot Number:

Performed By: 467497
Date: 4/9/2013

Documents

☒ Certificate
☒ Exception Message Guide
☒ Operator Guide
☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1472 MP1531

☒ Printer

Reviewed By: 079490
Date: 4/20/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/12/2013

Serial Number: 90-000280

AssetTag: 775-22071

Facility

Colorado State Patrol

Alamosa

Ship To:

Colorado State Patrol

3110 First Street

Alamosa, CO 81101

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99955 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/12/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1026 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1013 g/210L

☒ Standard Deviation (<0.003) .0006

☒ Relative Standard Deviation (<3%) .0063

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0280130001I

Performed By: 467497 Date: 4/12/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/12/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1318

MP1344

☒ Printer

Reviewed By: 079490

Date: 4/20/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/12/2013

Serial Number: 90-000281

AssetTag: 775-22072

Facility

Colorado State Patrol
Ship To:
Colorado State Patrol
1480 Quail Lake Loop
Colorado Springs, CO 80906

El Paso County

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99950 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/12/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0992 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1020 g/210L

☒ Standard Deviation (<0.003) .0006

☒ Relative Standard Deviation (<3%) .0064

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0281130001I

Performed By: 467497 Date: 4/12/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/12/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1310

MP1382

☒ Printer

Reviewed By: 079490

Date: 4/19/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/23/2013

Serial Number: 90-000282

AssetTag: 775-22073

Facility

Colorado State Patrol
Ship To:
Colorado State Patrol
3832 South I-25
Fort Collins, CO 80525

Fort Collins

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99961 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/23/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1005 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1005 g/210L

☒ Standard Deviation (<0.003) .0008

☒ Relative Standard Deviation (<3%) .0076

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0282130001I

Performed By: 732516 Date: 4/23/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/23/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1366

MP1577

☒ Printer

Reviewed By: 079490

Date: 4/23/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/9/2013
Serial Number: 90-000283
AssetTag: 775-22074

Facility

Colorado State Patrol
Ship To:
Colorado State Patrol
131 C Avenue
Limon, CO 80828

Limon

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99912 Pass ☒
Calibration Adjustment Pass ☒
Maintenance Log Entry ☒

Performed By: 467497 Date: 4/9/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0.0990 g/210L
☒ Linearity Test ☒ Interferant Detected
☒ Stability Test Average: 0.0979 g/210L
☒ Standard Deviation (<0.003) .0006
☒ Relative Standard Deviation (<3%) .0060
☒ EBAT Breath Test ☒ Final EBAT Test
☒ Improper Sample Test ☒ Maintenance Log Entry
☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com
☒ Card Reader ☒ Signature Pad
☒ IP Settings ☒ Modem
☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0283130001I

Performed By: 467497 Date: 4/9/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)
☒ Keyboard
☒ Signature Pad with USB Adapter
☒ Simulator Communication Cable
☒ Printer Communication Cable
☒ Trainer Access Card
☒ Power Supply Cables (3)
☒ Ethernet Cable
☒ Standard Simulator Solution

Lot Number:

Performed By: 467497
Date: 4/9/2013

Documents

☒ Certificate
☒ Exception Message Guide
☒ Operator Guide
☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1408 MP1499

☒ Printer

Reviewed By: 079490
Date: 4/17/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/26/2013

Serial Number: 90-000284

AssetTag: 775-22075

Facility

Colorado Springs Police Department

Stetson Hills

Ship To:

Colorado Springs Police Department

4110 Tutt Boulevard

Colorado Springs, CO 80922

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99969 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 079490 Date: 4/26/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0981 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0984 g/210L

☒ Standard Deviation (<0.003) .0007

☒ Relative Standard Deviation (<3%) .0068

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0284130001I

Performed By: 079490 Date: 4/26/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 079490

Date: 4/26/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1373

MP1582

☒ Printer

Reviewed By: 467497

Date: 4/26/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/26/2013

Serial Number: 90-000285

AssetTag: 775-22076

Facility

Eaton Police Department

Ship To:

Eaton Police Department

224 1st Street

Eaton, CO 80615

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99982 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 079490 Date: 4/26/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0981 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0972 g/210L

☒ Standard Deviation (<0.003) .0008

☒ Relative Standard Deviation (<3%) .0086

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0285130001I

Performed By: 079490 Date: 4/26/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 079490

Date: 4/26/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1383

MP1585

☒ Printer

Reviewed By: 467497

Date: 4/26/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/9/2013
Serial Number: 90-000286
AssetTag: 775-22077

Facility

CSU Police Department
Ship To:
CSU Police Department
750 Meridian Avenue
Fort Collins, CO 80523-6023

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99893 Pass ☒
Calibration Adjustment Pass ☒
Maintenance Log Entry ☒

Performed By: 467497 Date: 4/10/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0981 g/210L
☒ Linearity Test ☒ Interferant Detected
☒ Stability Test Average: 0.0969 g/210L
☒ Standard Deviation (<0.003) 0.0005
☒ Relative Standard Deviation (<3%) 0.0051
☒ EBAT Breath Test ☒ Final EBAT Test
☒ Improper Sample Test ☒ Maintenance Log Entry
☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com
☒ Card Reader ☒ Signature Pad
☒ IP Settings ☒ Modem
☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0286130001I

Performed By: 467497 Date: 4/10/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)
☒ Keyboard
☒ Signature Pad with USB Adapter
☒ Simulator Communication Cable
☒ Printer Communication Cable
☒ Trainer Access Card
☒ Power Supply Cables (3)
☒ Ethernet Cable
☒ Standard Simulator Solution

Lot Number:

Performed By: 467497
Date: 4/10/2013

Documents

☒ Certificate
☒ Exception Message Guide
☒ Operator Guide
☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1320 MP1330

☒ Printer

Reviewed By: 079490
Date: 4/17/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/23/2013

Serial Number: 90-000287

AssetTag: 775-22078

Facility

Commerce City Police Department

Ship To:

Commerce City Police Department

7887 E. 60th Avenue

Commerce City, CO 80022

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99896 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/23/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0987 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0982 g/210L

☒ Standard Deviation (<0.003) .0006

☒ Relative Standard Deviation (<3%) .0060

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0280130001I

Performed By: 732516 Date: 4/23/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/23/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1325

MP1363

☒ Printer

Reviewed By: 079490

Date: 4/23/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/12/2013

Serial Number: 90-000288

AssetTag: 775-22079

Facility

Conejos County Sheriff's Office

Ship To:

Conejos County Sheriff's Office

14044 County Road G5

Antonito, CO 81120

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99903 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/12/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0979 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0988 g/210L

☒ Standard Deviation (<0.003) .0008

☒ Relative Standard Deviation (<3%) .0084

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0288130001I

Performed By: 732516 Date: 4/12/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/12/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1599

MP1563

☒ Printer

Reviewed By: 079490

Date: 4/12/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/11/2013

Serial Number: 90-000290

AssetTag: 775-22081

Facility

Costilla County Sheriff's Office

Ship To:

Costilla County Sheriff's Office

103 Gaspar Street

San Luis, CO 81152

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99921 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/11/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0965 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0965 g/210L

☒ Standard Deviation (<0.003) .0006

☒ Relative Standard Deviation (<3%) .0063

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0290130001I

Performed By: 467497 Date: 4/11/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/11/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1556

MP1627

☒ Printer

Reviewed By: 079490

Date: 4/16/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/25/2013

Serial Number: 90-000291

AssetTag: 775-22082

Facility

Crested Butte Marshal's Dept

Ship To:

Crested Butte Marshal's Dept

508 Maroon Avenue

Crested Butte, CO 81224

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99953 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/25/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1002 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1000 g/210L

☒ Standard Deviation (<0.003) .0007

☒ Relative Standard Deviation (<3%) .0073

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0291130001I

Performed By: 732516 Date: 4/25/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/25/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1602

MP1661

☒ Printer

Reviewed By: 079490

Date: 4/26/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/26/2013

Serial Number: 90-000292

AssetTag: 775-22083

Facility

Cripple Creek Police Department

Ship To:

Cripple Creek Police Department

101 East Bennett Avenue

Cripple Creek, CO 80813

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99816 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/26/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0983 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0982 g/210L

☒ Standard Deviation (<0.003) .0006

☒ Relative Standard Deviation (<3%) .0060

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0292130001I

Performed By: 732516 Date: 4/26/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/26/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1604

MP1524

☒ Printer

Reviewed By: 079490

Date: 4/26/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/8/2013
Serial Number: 90-000293
AssetTag: 775-22084

Facility

Crowley County Sheriff's Office
Ship To:
Crowley County Sheriff's Office
6th and Main
Ordway, CO 81063

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99985 Pass ☒
Calibration Adjustment Pass ☒
Maintenance Log Entry ☒

Performed By: 467497 Date: 4/8/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1032 g/210L
☒ Linearity Test ☒ Interferant Detected
☒ Stability Test Average: 0.1006 g/210L
☒ Standard Deviation (<0.003) .0008
☒ Relative Standard Deviation (<3%) .0075
☒ EBAT Breath Test ☒ Final EBAT Test
☒ Improper Sample Test ☒ Maintenance Log Entry
☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com
☒ Card Reader ☒ Signature Pad
☒ IP Settings ☒ Modem
☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0293130001I

Performed By: 467497 Date: 4/9/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)
☒ Keyboard
☒ Signature Pad with USB Adapter
☒ Simulator Communication Cable
☒ Printer Communication Cable
☒ Trainer Access Card
☒ Power Supply Cables (3)
☒ Ethernet Cable
☒ Standard Simulator Solution

Lot Number:

Performed By: 467497
Date: 4/9/2013

Documents

☒ Certificate
☒ Exception Message Guide
☒ Operator Guide
☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1677 MP1548

☒ Printer

Reviewed By: 079490
Date: 4/17/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/10/2013

Serial Number: 90-000294

AssetTag: 775-22085

Facility

Delta County Sheriff's Office

Ship To:

Delta County Sheriff's Office

555 Palmer Street

Delta, CO 81416

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99962 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/10/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1005 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1002 g/210L

☒ Standard Deviation (<0.003) .0009

☒ Relative Standard Deviation (<3%) .0089

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0294130001I

Performed By: 467497 Date: 4/10/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/10/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1438

MP1544

☒ Printer

Reviewed By: 079490

Date: 4/20/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/10/2013

Serial Number: 90-000295

AssetTag: 775-22086

Facility

Delta County Sheriff

Hotchkiss

MAC Address:

Ship To:

IP Address:

Delta County Sheriff

Gateway:

555 Palmer Street

Subnet Mask:

Delta, CO 81416

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99947 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/10/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0981 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0978 g/210L

☒ Standard Deviation (<0.003) 0.0005

☒ Relative Standard Deviation (<3%) 0.0053

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0295130001I

Performed By: 467497 Date: 4/11/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/11/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1411

MP1427

☒ Printer

Reviewed By: 079490

Date: 4/16/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/26/2013

Serial Number: 90-000296

AssetTag: 775-22087

Facility

Denver Police Department

Traffic

Ship To:

Denver Police Department

3381 Park Avenue West

Denver, CO 80216

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.0990 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/26/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0990 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0994 g/210L

☒ Standard Deviation (<0.003) 0.0007

☒ Relative Standard Deviation (<3%) 0.0068

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0296130001I

Performed By: 732516 Date: 4/26/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/26/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

☒ Printer

Reviewed By: 079490

Date: 4/26/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/12/2013

Serial Number: 90-000297

AssetTag: 775-22088

Facility

Dolores County Sheriff's Office

Ship To:

Dolores County Sheriff's Office

409 North Main

Dove Creek, CO 81324-0505

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99964 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/12/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1004 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1014 g/210L

☒ Standard Deviation (<0.003) .0007

☒ Relative Standard Deviation (<3%) .0066

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0297130001I

Performed By: 467497 Date: 4/12/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/12/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1444

MP1520

☒ Printer

Reviewed By: 079490

Date: 4/26/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/25/2013

Serial Number: 90-000298

AssetTag: 775-22089

Facility

Douglas County Sheriff's Office

Jail

Ship To:

Douglas County Sheriff's Office

4000 Justice Way

Castle Rock, CO 80109

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99804 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/25/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0961 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0960 g/210L

☒ Standard Deviation (<0.003) 0.0007

☒ Relative Standard Deviation (<3%) 0.0072

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0289130001I

Performed By: 732516 Date: 4/25/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/25/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1319

MP1345

☒ Printer

Reviewed By: 467497

Date: 4/26/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/22/2013

Serial Number: 90-000299

AssetTag: 775-22090

Facility

Douglas County Sheriff's Office

Highlands Ranch

Ship To:

Douglas County Sheriff's Office

4000 Justice Way

Castle Rock, CO 80109

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99855 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/22/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1013 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.10103 g/210L

☒ Standard Deviation (<0.003) .0006

☒ Relative Standard Deviation (<3%) .0063

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

02999130001I

Performed By: 732516 Date: 4/22/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/22/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1638

MP1597

☒ Printer

Reviewed By: 079490

Date: 4/23/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/16/2013

Serial Number: 90-000301

AssetTag: 775-22092

Facility

Eagle County Sheriff's Office

El Jebel

Ship To:

Eagle County Sheriff's Office

20 Eagle County Drive

El Jebel, CO 81623

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99943 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 079490 Date: 4/17/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1013 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1005 g/210L

☒ Standard Deviation (<0.003) .0007

☒ Relative Standard Deviation (<3%) .0068

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0301130001I

Performed By: 079490 Date: 4/17/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 079490

Date: 4/17/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1504

MP1353

☒ Printer

Reviewed By: 467497

Date: 4/20/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/29/2013

Serial Number: 90-000302

AssetTag: 775-22093

Facility

Edgewater Police Department

Ship To:

Edgewater Police Department

5901 West 25th Avenue

Edgewater, CO 80214

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99958 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/29/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0998 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0999 g/210L

☒ Standard Deviation (<0.003) .0006

☒ Relative Standard Deviation (<3%) .0064

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0302130001I

Performed By: 732516 Date: 4/29/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/29/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1514

MP1583

☒ Printer

Reviewed By: 079490

Date: 4/29/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/17/2013

Serial Number: 90-000303

AssetTag: 775-22094

Facility

El Paso County Sheriff's Office

Ship To:

El Paso County Sheriff's Office

27 E. Vermijo Avenue

Colorado Springs, CO 80903

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) .99845 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/17/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0996 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0996 g/210L

☒ Standard Deviation (<0.003) .0007

☒ Relative Standard Deviation (<3%) .0068

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0303130001I

Performed By: 467497 Date: 4/17/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/17/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1595

MP1347

☒ Printer

Reviewed By: 079490

Date: 4/23/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/22/2013

Serial Number: 90-000304

AssetTag: 775-22095

Facility

Elbert County Sheriff's Office

Ship To:

Elbert County Sheriff's Office

751 Ute Avenue

Kiowa, CO 80117

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99942 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/22/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1010 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1010 g/210L

☒ Standard Deviation (<0.003) .0007

☒ Relative Standard Deviation (<3%) .0068

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0304130001I

Performed By: 732516 Date: 4/22/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/22/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1593

MP1453

☒ Printer

Reviewed By: 079490

Date: 4/23/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/24/2013

Serial Number: 90-000305

AssetTag: 775-22096

Facility

Elizabeth Police Department

Ship To:

Elizabeth Police Department

425 Main Street

Elizabeth, CO 80107-1527

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99897 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/24/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0991 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0979 g/210L

☒ Standard Deviation (<0.003) .0009

☒ Relative Standard Deviation (<3%) .0089

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0305130001I

Performed By: 732516 Date: 4/24/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/24/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1372

MP1500

☒ Printer

Reviewed By: 079490

Date: 4/25/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/9/2013
Serial Number: 90-000306
AssetTag: 775-22097

Facility

Englewood Police Department
Ship To:
Englewood Police Department
3615 South Elati Street
Englewood, CO 80110

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99984 Pass ☒
Calibration Adjustment Pass ☒
Maintenance Log Entry ☒

Performed By: 467497 Date: 4/9/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1028 g/210L
☒ Linearity Test ☒ Interferant Detected
☒ Stability Test Average: 0.1002 g/210L
☒ Standard Deviation (<0.003) .0005
☒ Relative Standard Deviation (<3%) .0052
☒ EBAT Breath Test ☒ Final EBAT Test
☒ Improper Sample Test ☒ Maintenance Log Entry
☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com
☒ Card Reader ☒ Signature Pad
☒ IP Settings ☒ Modem
☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0306130001I

Performed By: 467497 Date: 4/9/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)
☒ Keyboard
☒ Signature Pad with USB Adapter
☒ Simulator Communication Cable
☒ Printer Communication Cable
☒ Trainer Access Card
☒ Power Supply Cables (3)
☒ Ethernet Cable
☒ Standard Simulator Solution

Lot Number:

Performed By: 467497
Date: 4/9/2013

Documents

☒ Certificate
☒ Exception Message Guide
☒ Operator Guide
☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1401 MP1421

☒ Printer

Reviewed By: 079490
Date: 4/17/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/10/2013

Serial Number: 90-000307

AssetTag: 775-22098

Facility

Erie Police Department

Ship To:

Erie Police Department

645 Holbrook Street

Erie, CO 80516

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99900 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/10/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0981 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0971 g/210L

☒ Standard Deviation (<0.003) .0007

☒ Relative Standard Deviation (<3%) .0071

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0307130001I

Performed By: 467497 Date: 4/10/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/10/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1473

MP1536

☒ Printer

Reviewed By: 079490

Date: 4/20/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/26/2013

Serial Number: 90-000308

AssetTag: 775-22099

Facility

Estes Park Police Department

Ship To:

Estes Park Police Department

170 MacGregor Avenue

Estes Park, CO 80517

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99961 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/26/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0982 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0980 g/210L

☒ Standard Deviation (<0.003) .0006

☒ Relative Standard Deviation (<3%) .0057

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0308130001I

Performed By: 732516 Date: 4/26/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/26/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1641

MP1327

☒ Printer

Reviewed By: 079490

Date: 4/26/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/26/2013

Serial Number: 90-000310

AssetTag: 775-22101

Facility

Fort Carson Police Department

Ship To:

Fort Carson Police Department

7227 Christie Street, Bldg 2700

Fort Carson, CO 80913

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99973 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 079490 Date: 4/26/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0973 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0965 g/210L

☒ Standard Deviation (<0.003) .0006

☒ Relative Standard Deviation (<3%) .0063

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0310130001I

Performed By: 079490 Date: 4/26/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 079490

Date: 4/26/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1374

MP1668

☒ Printer

Reviewed By: 467497

Date: 4/26/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/22/2013

Serial Number: 90-000311

AssetTag: 775-22102

Facility

Fort Collins Police Department

Ship To:

Fort Collins Police Department

2221 S. Timberline

Fort Collins, CO 80525

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99935 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/22/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0982 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0984 g/210L

☒ Standard Deviation (<0.003) .0006

☒ Relative Standard Deviation (<3%) .0061

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0311130001I

Performed By: 732516 Date: 4/22/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/22/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1332

MP1397

☒ Printer

Reviewed By: 079490

Date: 4/23/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 5/13/2013

Serial Number: 90-000312

AssetTag: 775-22103

Facility

Fort Lupton Police Department

Ship To:

Fort Lupton Police Department

130 South McKinley Avenue

Fort Lupton, CO 80621

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99887 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 5/13/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0987 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0984 g/210L

☒ Standard Deviation (<0.003) 0.0008

☒ Relative Standard Deviation (<3%) 0.0083

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0312130001I

Performed By: 732516 Date: 5/13/2013

I-9000 Equipment Inventory

☐ Simulator Power Supplies (2)

☐ Keyboard

☐ Signature Pad with USB Adapter

☐ Simulator Communication Cable

☐ Printer Communication Cable

☐ Trainer Access Card

☐ Power Supply Cables (3)

☐ Ethernet Cable

☐ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 5/13/2013

Documents

☒ Certificate

☐ Exception Message Guide

☐ Operator Guide

☐ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

☐ Printer

Reviewed By: 467497

Date: 5/13/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/8/2013
Serial Number: 90-000314
AssetTag: 775-22105

Facility

Fraser/Winter Park Police Department
Ship To:
Fraser/Winter Park Police Department
79050 Main Street, Suite #4
Winter Park, CO 80482

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99838 Pass ☒
Calibration Adjustment Pass ☒
Maintenance Log Entry ☒

Performed By: 467497 Date: 4/8/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1009 g/210L
☒ Linearity Test ☒ Interferant Detected
☒ Stability Test Average: 0.0999 g/210L
☒ Standard Deviation (<0.003) .0010
☒ Relative Standard Deviation (<3%) .0097
☒ EBAT Breath Test ☒ Final EBAT Test
☒ Improper Sample Test ☒ Maintenance Log Entry
☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com
☒ Card Reader ☒ Signature Pad
☒ IP Settings ☒ Modem
☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0314130001I

Performed By: 467497 Date: 4/9/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)
☒ Keyboard
☒ Signature Pad with USB Adapter
☒ Simulator Communication Cable
☒ Printer Communication Cable
☒ Trainer Access Card
☒ Power Supply Cables (3)
☒ Ethernet Cable
☒ Standard Simulator Solution

Lot Number:

Performed By: 467497
Date: 4/9/2013

Documents

☒ Certificate
☒ Exception Message Guide
☒ Operator Guide
☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1567 MP1474

☒ Printer

Reviewed By: 079490
Date: 4/17/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/10/2013

Serial Number: 90-000315

AssetTag: 775-22106

Facility

Fremont County Sheriff's Office

Ship To:

Fremont County Sheriff's Office

100 Justice Center Road

Canon City, CO 81212

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99956 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/10/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0971 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0968 g/210L

☒ Standard Deviation (<0.003) .0011

☒ Relative Standard Deviation (<3%) .0111

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0315130001I

Performed By: 467497 Date: 4/10/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/10/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1534

MP1429

☒ Printer

Reviewed By: 079490

Date: 4/20/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/23/2013

Serial Number: 90-000316

AssetTag: 775-22107

Facility

Fruita Police Department
Ship To:
Fruita Police Department
101 West McCune Avenue
Fruita, CO 81521

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99973 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/23/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0998 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0991 g/210L

☒ Standard Deviation (<0.003) .0008

☒ Relative Standard Deviation (<3%) .0077

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0316130001I

Performed By: 732516 Date: 4/23/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/23/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1672

MP1495

☒ Printer

Reviewed By: 467497

Date: 4/23/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/23/2013

Serial Number: 90-000317

AssetTag: 775-22108

Facility

Garfield County Sheriff's Office

Ship To:

Garfield County Sheriff's Office

107 8th Street

Glenwood Springs, CO 81601

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99981 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/23/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0986 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0980 g/210L

☒ Standard Deviation (<0.003) .0009

☒ Relative Standard Deviation (<3%) .0088

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0317130001I

Performed By: 732516 Date: 4/23/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/23/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1439

MP1664

☒ Printer

Reviewed By: 079490

Date: 4/23/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/23/2013

Serial Number: 90-000318

AssetTag: 775-22109

Facility

Gilpin County Sheriff's Office

Ship To:

Gilpin County Sheriff's Office

2960 Dory Hill Road #300

Golden, CO 80422

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99888 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/23/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0969 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0964 g/210L

☒ Standard Deviation (<0.003) .0007

☒ Relative Standard Deviation (<3%) .0070

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0318130001I

Performed By: 732516 Date: 4/23/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/23/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1639

MP1392

☒ Printer

Reviewed By: 079490

Date: 4/23/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/9/2013
Serial Number: 90-000319
AssetTag: 775-22110

Facility

Glenwood Springs Police Department
Ship To:
Glenwood Springs Police Department
101 West 8th Street
Glenwood Springs, CO 81602

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99886 Pass ☒
Calibration Adjustment Pass ☒
Maintenance Log Entry ☒

Performed By: 467497 Date: 4/9/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1004 g/210L
☒ Linearity Test ☒ Interferant Detected
☒ Stability Test Average: 0.0999 g/210L
☒ Standard Deviation (<0.003) .0004
☒ Relative Standard Deviation (<3%) .0045
☒ EBAT Breath Test ☒ Final EBAT Test
☒ Improper Sample Test ☒ Maintenance Log Entry
☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com
☒ Card Reader ☒ Signature Pad
☒ IP Settings ☒ Modem
☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0319130001I

Performed By: 467497 Date: 4/9/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)
☒ Keyboard
☒ Signature Pad with USB Adapter
☒ Simulator Communication Cable
☒ Printer Communication Cable
☒ Trainer Access Card
☒ Power Supply Cables (3)
☒ Ethernet Cable
☒ Standard Simulator Solution

Lot Number:

Performed By: 467497
Date: 4/9/2013

Documents

☒ Certificate
☒ Exception Message Guide
☒ Operator Guide
☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1537 MP1406

☒ Printer

Reviewed By: 079490
Date: 4/20/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/16/2013

Serial Number: 90-000320

AssetTag: 775-22111

Facility

Golden Police Department

Ship To:

Golden Police Department

911 10th Street

Golden, CO 80401

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99964 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/16/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1004 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1003 g/210L

☒ Standard Deviation (<0.003) .0009

☒ Relative Standard Deviation (<3%) .0085

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0320130001I

Performed By: 467497 Date: 4/17/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/17/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1687

MP1557

☒ Printer

Reviewed By: 079490

Date: 4/17/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/9/2013
Serial Number: 90-000321
AssetTag: 775-22112

Facility

Grand County Sheriff's Office
Ship To:
Grand County Sheriff's Office
670 Spring Street
Hot Sulphur Springs, CO 80451

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99995 Pass ☒
Calibration Adjustment Pass ☒
Maintenance Log Entry ☒

Performed By: 467497 Date: 4/9/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0988 g/210L
☒ Linearity Test ☒ Interferant Detected
☒ Stability Test Average: 0.0983 g/210L
☒ Standard Deviation (<0.003) .0013
☒ Relative Standard Deviation (<3%) .0127
☒ EBAT Breath Test ☒ Final EBAT Test
☒ Improper Sample Test ☒ Maintenance Log Entry
☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com
☒ Card Reader ☒ Signature Pad
☒ IP Settings ☒ Modem
☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0321130001I

Performed By: 467497 Date: 4/9/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)
☒ Keyboard
☒ Signature Pad with USB Adapter
☒ Simulator Communication Cable
☒ Printer Communication Cable
☒ Trainer Access Card
☒ Power Supply Cables (3)
☒ Ethernet Cable
☒ Standard Simulator Solution

Lot Number:

Performed By: 467497
Date: 4/9/2013

Documents

☒ Certificate
☒ Exception Message Guide
☒ Operator Guide
☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1398 MP1671

☒ Printer

Reviewed By: 079490
Date: 4/17/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/23/2013

Serial Number: 90-000322

AssetTag: 775-22113

Facility

Lakewood Police Department

Ship To:

Lakewood Police Department

445 South Allison Parkway

Lakewood, CO 80226

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99949 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/23/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0997 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.999 g/210L

☒ Standard Deviation (<0.003) .0009

☒ Relative Standard Deviation (<3%) .0085

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0322130001I

Performed By: 732516 Date: 4/23/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/23/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1352

MP1589

☒ Printer

Reviewed By: 079490

Date: 4/25/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/17/2013

Serial Number: 90-000324

AssetTag: 775-22115

Facility

Gunnison County Sheriff's Office

Ship To:

Gunnison County Sheriff's Office

201 West Virginia Avenue

Gunnison, CO 81230

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99995 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/17/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0992 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0993 g/210L

☒ Standard Deviation (<0.003) .0007

☒ Relative Standard Deviation (<3%) .0066

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0324130001I

Performed By: 732516 Date: 4/17/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/17/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1475

MP1541

☒ Printer

Reviewed By: 079490

Date: 4/20/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/4/2013
Serial Number: 90-000325
AssetTag: 775-22116

Facility

Huerfano County Sheriff's Office
Ship To:
Huerfano County Sheriff's Office
500 South Albert Street
Walsenburg, CO 81089

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99966 Pass ☒
Calibration Adjustment Pass ☒
Maintenance Log Entry ☒

Performed By: 467497 Date: 4/4/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0.1014 g/210L
☒ Linearity Test ☒ Interferant Detected
☒ Stability Test Average: 0.1016 g/210L
☒ Standard Deviation (<0.003) .0005
☒ Relative Standard Deviation (<3%) .0050
☒ EBAT Breath Test ☒ Final EBAT Test
☒ Improper Sample Test ☒ Maintenance Log Entry
☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com
☒ Card Reader ☒ Signature Pad
☒ IP Settings ☒ Modem
☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0325130001I

Performed By: 467497 Date: 4/5/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)
☒ Keyboard
☒ Signature Pad with USB Adapter
☒ Simulator Communication Cable
☒ Printer Communication Cable
☒ Trainer Access Card
☒ Power Supply Cables (3)
☒ Ethernet Cable
☒ Standard Simulator Solution

Lot Number:

Performed By: 467497
Date: 4/5/2013

Documents

☒ Certificate
☒ Exception Message Guide
☒ Operator Guide
☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1611 MP1377

☒ Printer

Reviewed By: 079490
Date: 4/16/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 5/13/2013

Serial Number: 90-000326

AssetTag: 775-22117

Facility

Jackson County Sheriff's Office

Ship To:

Jackson County Sheriff's Office

396 La Fever

Walden, CO 80480

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99966 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 5/13/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0995 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0997 g/210L

☒ Standard Deviation (<0.003) 0.0009

☒ Relative Standard Deviation (<3%) 0.0093

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0326130001I

Performed By: 732516 Date: 5/13/2013

I-9000 Equipment Inventory

☐ Simulator Power Supplies (2)

☐ Keyboard

☐ Signature Pad with USB Adapter

☐ Simulator Communication Cable

☐ Printer Communication Cable

☐ Trainer Access Card

☐ Power Supply Cables (3)

☐ Ethernet Cable

☐ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 5/13/2013

Documents

☒ Certificate

☐ Exception Message Guide

☐ Operator Guide

☐ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

☐ Printer

Reviewed By: 467497

Date: 5/20/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/17/2013

Serial Number: 90-000327

AssetTag: 775-22118

Facility

Jefferson County Sheriff's Office

Ship To:

Jefferson County Sheriff's Office

200 Jefferson County Parkway

Golden, CO 80401

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99987 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/17/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0988 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0988 g/210L

☒ Standard Deviation (<0.003) .0008

☒ Relative Standard Deviation (<3%) .0078

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0327130001I

Performed By: 467497 Date: 4/17/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/17/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1501

MP1679

☒ Printer

Reviewed By: 079490

Date: 4/17/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/24/2013

Serial Number: 90-000328

AssetTag: 775-22119

Facility

Kit Carson County Sheriff's Office

Ship To:

Kit Carson County Sheriff's Office

251 16th Street

Burlington, CO 80807

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99937 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/24/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0971 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0973 g/210L

☒ Standard Deviation (<0.003) .0007

☒ Relative Standard Deviation (<3%) .0068

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0328130001I

Performed By: 732516 Date: 4/24/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/24/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1322

MP1338

☒ Printer

Reviewed By: 079490

Date: 4/25/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/9/2013
Serial Number: 90-000329
AssetTag: 775-22120

Facility

Lafayette Police Department
Ship To:
Lafayette Police Department
451 N. 111th Street
Laffayette, CO 80026

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99935 Pass ☒
Calibration Adjustment Pass ☒
Maintenance Log Entry ☒

Performed By: 467497 Date: 4/9/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0992 g/210L
☒ Linearity Test ☒ Interferant Detected
☒ Stability Test Average: 0.1002 g/210L
☒ Standard Deviation (<0.003) .0007
☒ Relative Standard Deviation (<3%) .0067
☒ EBAT Breath Test ☒ Final EBAT Test
☒ Improper Sample Test ☒ Maintenance Log Entry
☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com
☒ Card Reader ☒ Signature Pad
☒ IP Settings ☒ Modem
☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0329130001I

Performed By: 467497 Date: 4/9/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)
☒ Keyboard
☒ Signature Pad with USB Adapter
☒ Simulator Communication Cable
☒ Printer Communication Cable
☒ Trainer Access Card
☒ Power Supply Cables (3)
☒ Ethernet Cable
☒ Standard Simulator Solution

Lot Number:

Performed By: 467497
Date: 4/9/2013

Documents

☒ Certificate
☒ Exception Message Guide
☒ Operator Guide
☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1478 MP1433

☒ Printer

Reviewed By: 079490
Date: 4/9/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/23/2013

Serial Number: 90-000330

AssetTag: 775-22121

Facility

Greeley Police Department

Ship To:

Greeley Police Department

2875 West 10th Street

Greeley, CO 80634

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99910 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/23/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1009 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1010 g/210L

☒ Standard Deviation (<0.003) .0007

☒ Relative Standard Deviation (<3%) .0068

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0330130001I

Performed By: 732516 Date: 4/23/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/23/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1458

MP1608

☒ Printer

Reviewed By: 079490

Date: 4/25/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/29/2013

Serial Number: 90-000332

AssetTag: 775-22123

Facility

Las Animas County Sheriff's Office

Ship To:

Las Animas County Sheriff's Office

2309 East Main Street

Trinidad, CO 81082

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) .99995 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/29/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0996 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0996 g/210L

☒ Standard Deviation (<0.003) 0.0008

☒ Relative Standard Deviation (<3%) 0.0076

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0332130001I

Performed By: 732516 Date: 4/26/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/26/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

☒ Printer

Reviewed By: 467497

Date: 4/29/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/9/2013
Serial Number: 90-000333
AssetTag: 775-22124

Facility

Lincoln County Sheriff's Office
Ship To:
Lincoln County Sheriff's Office
103 3rd Avenue
Hugo, CO 80821

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99887 Pass ☒
Calibration Adjustment Pass ☒
Maintenance Log Entry ☒

Performed By: 467497 Date: 4/10/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0996 g/210L
☒ Linearity Test ☒ Interferant Detected
☒ Stability Test Average: 0.0996 g/210L
☒ Standard Deviation (<0.003) 0.0007
☒ Relative Standard Deviation (<3%) 0.0069
☒ EBAT Breath Test ☒ Final EBAT Test
☒ Improper Sample Test ☒ Maintenance Log Entry
☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com
☒ Card Reader ☒ Signature Pad
☒ IP Settings ☒ Modem
☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0333130001I

Performed By: 467497 Date: 4/10/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)
☒ Keyboard
☒ Signature Pad with USB Adapter
☒ Simulator Communication Cable
☒ Printer Communication Cable
☒ Trainer Access Card
☒ Power Supply Cables (3)
☒ Ethernet Cable
☒ Standard Simulator Solution

Lot Number:

Performed By: 467497
Date: 4/10/2013

Documents

☒ Certificate
☒ Exception Message Guide
☒ Operator Guide
☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1431 MP1477

☒ Printer

Reviewed By: 079490
Date: 4/16/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/23/2013

Serial Number: 90-000334

AssetTag: 775-22125

Facility

Lake County Sheriff's Office

Ship To:

Lake County Sheriff's Office

500 Harrison Avenue

Leadville, CO 80461

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99909 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/23/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.093 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0975 g/210L

☒ Standard Deviation (<0.003) .0006

☒ Relative Standard Deviation (<3%) .0062

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0334130001I

Performed By: 732516 Date: 4/23/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/23/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1622

MP1628

☒ Printer

Reviewed By: 079490

Date: 4/23/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/4/2013
Serial Number: 90-000335
AssetTag: 775-22126

Facility

Littleton Police Department
Ship To:
Littleton Police Department
2255 West Berry Avenue
Littleton, CO 80165

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.98870 Pass ☒
Calibration Adjustment Pass ☒
Maintenance Log Entry ☒

Performed By: 467497 Date: 4/4/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1010 g/210L
☒ Linearity Test ☒ Interferant Detected
☒ Stability Test Average: .1007 g/210L
☒ Standard Deviation (<0.003) .0007
☒ Relative Standard Deviation (<3%) .0073
☒ EBAT Breath Test ☒ Final EBAT Test
☒ Improper Sample Test ☒ Maintenance Log Entry
☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com
☒ Card Reader ☒ Signature Pad
☒ IP Settings ☒ Modem
☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0335130002I

Performed By: 467497 Date: 4/5/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)
☒ Keyboard
☒ Signature Pad with USB Adapter
☒ Simulator Communication Cable
☒ Printer Communication Cable
☒ Trainer Access Card
☒ Power Supply Cables (3)
☒ Ethernet Cable
☒ Standard Simulator Solution

Lot Number:

Performed By: 467497
Date: 4/5/2013

Documents

☒ Certificate
☒ Exception Message Guide
☒ Operator Guide
☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1519 MP1405

☒ Printer

Reviewed By: 079490
Date: 4/16/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/12/2013

Serial Number: 90-000336

AssetTag: 775-22127

Facility

Logan County Sheriff's Office

Ship To:

Logan County Sheriff's Office

110 North Riverview Road, Room 116

Sterling, CO 80751

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99988 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/12/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1002 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1006 g/210L

☒ Standard Deviation (<0.003) 0.0008

☒ Relative Standard Deviation (<3%) 0.0075

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0336130001i

Performed By: 467497 Date: 4/12/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/12/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1579

MP1342

☒ Printer

Reviewed By: 079490

Date: 4/16/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/23/2013

Serial Number: 90-000337

AssetTag: 775-22128

Facility

Lone Tree Police Department

Ship To:

Lone Tree Police Department

9220 Kimmer Drive

Lone Tree, CO 80124

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99943 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/23/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0982 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0973 g/210L

☒ Standard Deviation (<0.003) .0006

☒ Relative Standard Deviation (<3%) .0059

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0337130001I

Performed By: 732516 Date: 4/23/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/23/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1446

MP1455

☒ Printer

Reviewed By: 079490

Date: 4/23/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/12/2013

Serial Number: 90-000338

AssetTag: 775-22129

Facility

Longmont Police Department

Ship To:

Longmont Police Department

225 Kimbark Street

Longmont, CO 80501

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99965 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/12/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0980 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0994 g/210L

☒ Standard Deviation (<0.003) .0007

☒ Relative Standard Deviation (<3%) .0073

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0338130001I

Performed By: 467497 Date: 4/12/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/12/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1580

MP1454

☒ Printer

Reviewed By: 079490

Date: 4/19/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/16/2013

Serial Number: 90-000339

AssetTag: 775-22130

Facility

Louisville Police Department

Ship To:

Louisville Police Department

992 Via Appia

Louisville, CO 80027

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99996 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 079490 Date: 4/16/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1022 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1017 g/210L

☒ Standard Deviation (<0.003) .0006

☒ Relative Standard Deviation (<3%) .0058

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0339130001I

Performed By: 079490 Date: 4/17/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 079490

Date: 4/17/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1355

MP1535

☒ Printer

Reviewed By: 467497

Date: 4/17/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/24/2013

Serial Number: 90-000341

AssetTag: 775-22132

Facility

Manitou Springs Police Department

Ship To:

Manitou Springs Police Department

606 Manitou Avenue

Manitou Springs, CO 80829

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99870 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/24/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0986 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1009 g/210L

☒ Standard Deviation (<0.003) .0009

☒ Relative Standard Deviation (<3%) .0093

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0341130001I

Performed By: 732516 Date: 4/24/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/24/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1560

MP1551

☒ Printer

Reviewed By: 079490

Date: 4/25/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/23/2013

Serial Number: 90-000342

AssetTag: 775-22133

Facility

Mesa County Sheriff's Office

Ship To:

Mesa County Sheriff's Office

215 Rice Street

Grand Junction, CO 81501

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99840 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/23/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0984 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0987 g/210L

☒ Standard Deviation (<0.003) .0009

☒ Relative Standard Deviation (<3%) .0089

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0342120001I

Performed By: 732516 Date: 4/23/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/23/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1505

MP1562

☒ Printer

Reviewed By: 079490

Date: 4/23/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/19/2013

Serial Number: 90-000343

AssetTag: 775-22134

Facility

Mesa County Sheriff's Office

Collbran

Ship To:

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99902 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/19/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0985 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0984 g/210L

☒ Standard Deviation (<0.003) .0009

☒ Relative Standard Deviation (<3%) .0089

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0343130001I

Performed By: 732516 Date: 4/19/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/19/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1625

MP1674

☒ Printer

Reviewed By: 079490

Date: 4/20/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/9/2013
Serial Number: 90-000344
AssetTag: 775-22135

Facility

Moffat County Sheriff's Office
Ship To:
Moffat County Sheriff's Office
800 West First Street, Suite 300
Craig, CO 81625

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99971 Pass ☒
Calibration Adjustment Pass ☒
Maintenance Log Entry ☒

Performed By: 467497 Date: 4/9/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1010 g/210L
☒ Linearity Test ☒ Interferant Detected
☒ Stability Test Average: 0.1015 g/210L
☒ Standard Deviation (<0.003) .0005
☒ Relative Standard Deviation (<3%) .0050
☒ EBAT Breath Test ☒ Final EBAT Test
☒ Improper Sample Test ☒ Maintenance Log Entry
☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com
☒ Card Reader ☒ Signature Pad
☒ IP Settings ☒ Modem
☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0344130001I

Performed By: 467497 Date: 4/9/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)
☒ Keyboard
☒ Signature Pad with USB Adapter
☒ Simulator Communication Cable
☒ Printer Communication Cable
☒ Trainer Access Card
☒ Power Supply Cables (3)
☒ Ethernet Cable
☒ Standard Simulator Solution

Lot Number:

Performed By: 467497
Date: 4/9/2013

Documents

☒ Certificate
☒ Exception Message Guide
☒ Operator Guide
☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1445 MP1631

☒ Printer

Reviewed By: 079490
Date: 4/19/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/12/2013

Serial Number: 90-000345

AssetTag: 775-22136

Facility

Monte Vista Police Department

Ship To:

Monte Vista Police Department

95 W. 1st Avenue

Monte Vista, CO 81144

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99996 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/12/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0995 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0991 g/210L

☒ Standard Deviation (<0.003) 0.0009

☒ Relative Standard Deviation (<3%) 0.0092

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0345130001I

Performed By: 467497 Date: 4/12/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/12/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1399

MP1549

☒ Printer

Reviewed By: 079490

Date: 4/16/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/22/2013

Serial Number: 90-000347

AssetTag: 775-22138

Facility

Montrose County Sheriff's Office

Nucla

Ship To:

Montrose County Sheriff's Office

27700 DD Road

Nucla, CO 81424

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99921 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/22/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0981 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0977 g/210L

☒ Standard Deviation (<0.003) .0007

☒ Relative Standard Deviation (<3%) .0067

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0347130001I

Performed By: 732516 Date: 4/22/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/22/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1497

MP1349

☒ Printer

Reviewed By: 079490

Date: 4/23/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/17/2013

Serial Number: 90-000348

AssetTag: 775-22139

Facility

Montrose Police Department

Ship To:

Montrose Police Department

434 South 1st Street

Montrose, CO 81401

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) .099919 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/17/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0989 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0991 g/210L

☒ Standard Deviation (<0.003) .0007

☒ Relative Standard Deviation (<3%) .0069

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0348130001I

Performed By: 467497 Date: 4/17/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/17/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1616

MP1407

☒ Printer

Reviewed By: 079490

Date: 4/18/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/26/2013

Serial Number: 90-000349

AssetTag: 775-22140

Facility

Monument Police Department

Ship To:

Monument Police Department

645 Beacon Lite Road

Mounument, CO 80132

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99949 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/26/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0974 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0973 g/210L

☒ Standard Deviation (<0.003) .0006

☒ Relative Standard Deviation (<3%) .0066

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0349130001I

Performed By: 732516 Date: 4/26/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/26/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1357

MP1388

☒ Printer

Reviewed By: 079490

Date: 4/26/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/23/2013

Serial Number: 90-000350

AssetTag: 775-22141

Facility

Morgan County Sheriff's Office

Ship To:

Morgan County Sheriff's Office

801 East Beaver Avenue

Fort Morgan, CO 80701

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99998 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/23/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0978 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0981 g/210L

☒ Standard Deviation (<0.003) .0006

☒ Relative Standard Deviation (<3%) .0065

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0350130001I

Performed By: 732516 Date: 4/23/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/23/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1673

MP1483

☒ Printer

Reviewed By: 079490

Date: 4/23/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/17/2013

Serial Number: 90-000351

AssetTag: 775-22142

Facility

Mountain Village Police Department

Ship To:

Mountain Village Police Department

411 Mountain Village Boulevard

Mountain Village, CO 81435

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99927 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/17/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0987 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0984 g/210L

☒ Standard Deviation (<0.003) .0007

☒ Relative Standard Deviation (<3%) .0069

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0351130001I

Performed By: 732516 Date: 4/17/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/17/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1394

MP1317

☒ Printer

Reviewed By: 079490

Date: 4/17/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/22/2013

Serial Number: 90-000352

AssetTag: 775-22143

Facility

Northglenn Police Department

Ship To:

Northglenn Police Department

11701 Community Center Drive

Northglenn, CO 80233

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99893 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/22/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0988 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0981 g/210L

☒ Standard Deviation (<0.003) .0007

☒ Relative Standard Deviation (<3%) .0073

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0352130001I

Performed By: 732516 Date: 4/22/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/22/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1527

MP1443

☒ Printer

Reviewed By: 079490

Date: 4/22/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/24/2013

Serial Number: 90-000354

AssetTag: 775-22145

Facility

Palisade Police Department

Ship To:

Palisade Police Department

175 East 3rd Street

Palisade, CO 81526

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99923 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/24/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0991 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0990 g/210L

☒ Standard Deviation (<0.003) .0008

☒ Relative Standard Deviation (<3%) .0077

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0354130001I

Performed By: 732516 Date: 4/24/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/24/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1675

MP1346

☒ Printer

Reviewed By: 079490

Date: 4/25/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/23/2013

Serial Number: 90-000355

AssetTag: 775-22146

Facility

Parachute Police Department

Ship To:

Parachute Police Department

222 Grand Valley Way

Parachute, CO 81635

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99966 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/23/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1015 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1010 g/210L

☒ Standard Deviation (<0.003) .0006

☒ Relative Standard Deviation (<3%) .0064

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0355130001I

Performed By: 732516 Date: 4/23/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/23/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1351

MP1637

☒ Printer

Reviewed By: 079490

Date: 4/23/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/26/2013

Serial Number: 90-000357

AssetTag: 775-22148

Facility

Park County Sheriff's Office

Fairplay

Ship To:

Park County Sheriff's Office

1180 PCR 16

Fairplay, CO 80440

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99976 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 079490 Date: 4/26/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0982 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0985 g/210L

☒ Standard Deviation (<0.003) .0008

☒ Relative Standard Deviation (<3%) .0077

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0357130001I

Performed By: 079490 Date: 4/26/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 079490

Date: 4/26/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1341

MP1410

☒ Printer

Reviewed By: 467497

Date: 4/26/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/26/2013

Serial Number: 90-000358

AssetTag: 775-22149

Facility

Parker Police Department
Ship To:
Parker Police Department
18600 Lincoln Meadows Parkway
Parker, CO 80134

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99931 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/26/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0982 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0950 g/210L

☒ Standard Deviation (<0.003) .0006

☒ Relative Standard Deviation (<3%) .0064

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0358130001I

Performed By: 732516 Date: 4/26/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/26/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1652

MP1558

☒ Printer

Reviewed By: 467497

Date: 4/29/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/25/2013

Serial Number: 90-000359

AssetTag: 775-22150

Facility

Phillips County Sheriff's Office

Ship To:

Phillips County Sheriff's Office

221 South Inter Ocean

Holyoke, CO 80734

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99993 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/25/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0978 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0975 g/210L

☒ Standard Deviation (<0.003) .0012

☒ Relative Standard Deviation (<3%) .0122

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0359130001I

Performed By: 732516 Date: 4/25/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/25/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1326

MP1645

☒ Printer

Reviewed By: 467497

Date: 4/25/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 5/13/2013

Serial Number: 90-000360

AssetTag: 775-22151

Facility

Pitkin County Sheriff's Office

Ship To:

Pitkin County Sheriff's Office

506 E. Main Street, Suite 101

Aspen, CO 81611-2923

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99851 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 5/13/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0991 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0989 g/210L

☒ Standard Deviation (<0.003) 0.0005

☒ Relative Standard Deviation (<3%) 0.0050

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0360130001I

Performed By: 732516 Date: 5/13/2013

I-9000 Equipment Inventory

☐ Simulator Power Supplies (2)

☐ Keyboard

☐ Signature Pad with USB Adapter

☐ Simulator Communication Cable

☐ Printer Communication Cable

☐ Trainer Access Card

☐ Power Supply Cables (3)

☐ Ethernet Cable

☐ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 5/13/2013

Documents

☒ Certificate

☐ Exception Message Guide

☐ Operator Guide

☐ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

☐ Printer

Reviewed By: 467497

Date: 5/20/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/16/2013

Serial Number: 90-000363

AssetTag: 775-22154

Facility

Pueblo Police Department

Ship To:

Pueblo Police Department

200 South Main Street

Pueblo, CO 81003

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99984 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 079490 Date: 4/16/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1006 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1007 g/210L

☒ Standard Deviation (<0.003) .0009

☒ Relative Standard Deviation (<3%) .0087

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0363130001I

Performed By: 079490 Date: 4/17/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 079490

Date: 4/17/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1437

MP1683

☒ Printer

Reviewed By: 467497

Date: 4/20/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/23/2013

Serial Number: 90-000365

AssetTag: 775-22156

Facility

Colorado Parks and Wildlife

Ridgway

Ship To:

Colorado Parks and Wildlife

28555 Highway 550

Ridgway, CO 81432

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99927 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/10/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0982 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0968 g/210L

☒ Standard Deviation (<0.003) 0.0008

☒ Relative Standard Deviation (<3%) 0.0084

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0365130001I

Performed By: 732516 Date: 4/23/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/23/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1615

MP1547

☒ Printer

Reviewed By: 079490

Date: 4/23/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/23/2013

Serial Number: 90-000366

AssetTag: 775-22157

Facility

Rifle Police Department

Ship To:

Rifle Police Department

201 E. 18th Street

Rifle, CO 81650

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99981 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/23/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1004 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1002 g/210L

☒ Standard Deviation (<0.003) .0006

☒ Relative Standard Deviation (<3%) .0059

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0366130001I

Performed By: 732516 Date: 4/23/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/23/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1432

MP1511

☒ Printer

Reviewed By: 079490

Date: 4/23/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 5/13/2013

Serial Number: 90-000367

AssetTag: 775-22158

Facility

Rio Grande County Sheriff's Office

Ship To:

Rio Grande County Sheriff's Office

640 Cherry Street

Del Norte, CO 81132

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99916 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 5/13/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0985 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0987 g/210L

☒ Standard Deviation (<0.003) 0.0007

☒ Relative Standard Deviation (<3%) 0.0074

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0367130001I

Performed By: 732516 Date: 5/13/2013

I-9000 Equipment Inventory

☐ Simulator Power Supplies (2)

☐ Keyboard

☐ Signature Pad with USB Adapter

☐ Simulator Communication Cable

☐ Printer Communication Cable

☐ Trainer Access Card

☐ Power Supply Cables (3)

☐ Ethernet Cable

☐ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 5/13/2013

Documents

☒ Certificate

☐ Exception Message Guide

☐ Operator Guide

☐ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

☐ Printer

Reviewed By: 467497

Date: 5/20/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/25/2013

Serial Number: 90-000368

AssetTag: 775-22159

Facility

Routt County Sheriff's Office
Ship To:
Routt County Sheriff's Office
2025 Shield Drive
Steamboat Springs, CO 80487

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99961 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/25/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0971 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0972 g/210L

☒ Standard Deviation (<0.003) .0007

☒ Relative Standard Deviation (<3%) .0072

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0368130001I

Performed By: 732516 Date: 4/25/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/25/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1513

MP1592

☒ Printer

Reviewed By: 079490

Date: 4/26/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/5/2013
Serial Number: 90-000370
AssetTag: 775-22161

Facility

Sheridan Police Department
Ship To:
Sheridan Police Department
4101 South Federal Boulevard
Sheridan, CO 80110

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99937 Pass ☒
Calibration Adjustment Pass ☒
Maintenance Log Entry ☒

Performed By: 467497 Date: 4/5/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1013 g/210L
☒ Linearity Test ☒ Interferant Detected
☒ Stability Test Average: 0.1010 g/210L
☒ Standard Deviation (<0.003) .0004
☒ Relative Standard Deviation (<3%) .0039
☒ EBAT Breath Test ☒ Final EBAT Test
☒ Improper Sample Test ☒ Maintenance Log Entry
☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com
☒ Card Reader ☒ Signature Pad
☒ IP Settings ☒ Modem
☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0370130001I

Performed By: 467497 Date: 4/5/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)
☒ Keyboard
☒ Signature Pad with USB Adapter
☒ Simulator Communication Cable
☒ Printer Communication Cable
☒ Trainer Access Card
☒ Power Supply Cables (3)
☒ Ethernet Cable
☒ Standard Simulator Solution

Lot Number:

Performed By: 467497
Date: 4/5/2013

Documents

☒ Certificate
☒ Exception Message Guide
☒ Operator Guide
☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1343 MP1448

☒ Printer

Reviewed By: 079490
Date: 4/20/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/17/2013

Serial Number: 90-000371

AssetTag: 775-22162

Facility

Silverthorne Police Department

Ship To:

Silverthorne Police Department

601 Center Circle

Silverthorne, CO 80498

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99878 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/17/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0994 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0990 g/210L

☒ Standard Deviation (<0.003) .0006

☒ Relative Standard Deviation (<3%) .0057

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0371130001I

Performed By: 467497 Date: 4/17/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/17/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1576

MP1379

☒ Printer

Reviewed By: 079490

Date: 4/20/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/16/2013

Serial Number: 90-000372

AssetTag: 775-22163

Facility

Snowmass Police Department

Ship To:

Snowmass Police Department

130 Kearns Road

Snowmass Village, CO 81615

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99961 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 079490 Date: 4/16/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1018 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1011 g/210L

☒ Standard Deviation (<0.003) .0008

☒ Relative Standard Deviation (<3%) .0082

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0372130001I

Performed By: 079490 Date: 4/17/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 079490

Date: 4/17/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1428

MP1449

☒ Printer

Reviewed By: 467497

Date: 4/20/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/22/2013

Serial Number: 90-000373

AssetTag: 775-22164

Facility

Southern Ute Tribal Police Department

Ship To:

Southern Ute Tribal Police Department

149 County Road 517

Ignacio, CO 81137

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99817 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/22/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1004 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1006 g/210L

☒ Standard Deviation (<0.003) .0006

☒ Relative Standard Deviation (<3%) .0060

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0373130001I

Performed By: 732516 Date: 4/22/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/22/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1603

MP1656

☒ Printer

Reviewed By: 079490

Date: 4/23/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/17/2013

Serial Number: 90-000374

AssetTag: 775-22165

Facility

Summit County Sheriff's Office

Ship To:

Summit County Sheriff's Office

501 North Park Avenue

Breckenridge, CO 80424

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99999 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/17/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0998 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1000 g/210L

☒ Standard Deviation (<0.003) .0005

☒ Relative Standard Deviation (<3%) .0051

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0374130001I

Performed By: 467497 Date: 4/17/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/17/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1487

MP1554

☒ Printer

Reviewed By: 079490

Date: 4/17/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/23/2013

Serial Number: 90-000375

AssetTag: 775-22166

Facility

Teller County Sheriff's Office

Ship To:

Teller County Sheriff's Office

11400 West Highway 24

Divide, CO 80814

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99983 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/25/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0989 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0987 g/210L

☒ Standard Deviation (<0.003) .0006

☒ Relative Standard Deviation (<3%) .0058

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0375130001I

Performed By: 732516 Date: 4/25/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/25/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1621

MP1682

☒ Printer

Reviewed By: 467497

Date: 4/26/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/25/2013

Serial Number: 90-000376

AssetTag: 775-22167

Facility

Thornton Police Department

Ship To:

Thornton Police Department

9551 Civic Center Drive

Thornton, CO 80229

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99989 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/25/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0958 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0952 g/210L

☒ Standard Deviation (<0.003) 0.0006

☒ Relative Standard Deviation (<3%) 0.0065

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0376130001I

Performed By: 732516 Date: 4/25/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/25/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1578

MP1365

☒ Printer

Reviewed By: 467497

Date: 4/26/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/10/2013

Serial Number: 90-000377

AssetTag: 775-22168

Facility

University of Colorado at Colorado Springs Police Department
Ship To:
University of Colorado at Colorado Springs Police
Department
1420 Austin Bluffs Parkway
Colorado Springs, CO 80918

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99857 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/10/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1002 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0999 g/210L

☒ Standard Deviation (<0.003) .0006

☒ Relative Standard Deviation (<3%) .0055

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

03771300011

Performed By: 467497 Date: 4/13/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/13/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1442

MP1540

☒ Printer

Reviewed By: 079490

Date: 4/16/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/17/2013

Serial Number: 90-000378

AssetTag: 775-22169

Facility

Vail Police Department
Ship To:
Vail Police Department
75 South Frontage Road
Vail, CO 81658

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99965 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/17/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0988 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0988 g/210L

☒ Standard Deviation (<0.003) .0009

☒ Relative Standard Deviation (<3%) .0086

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0378130001I

Performed By: 467497 Date: 4/18/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/18/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1542

MP1414

☒ Printer

Reviewed By: 079490

Date: 4/20/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/25/2013

Serial Number: 90-000379

AssetTag: 775-22170

Facility

Washington County Sheriff's Office

Ship To:

Washington County Sheriff's Office

26861 Highway 34

Akron, CO 80720

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99924 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/25/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0997 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0987 g/210L

☒ Standard Deviation (<0.003) .0009

☒ Relative Standard Deviation (<3%) .0094

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0379130001I

Performed By: 732516 Date: 4/25/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/25/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1658

MP1662

☒ Printer

Reviewed By: 079490

Date: 4/26/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/25/2013

Serial Number: 90-000380

AssetTag: 775-22171

Facility

Weld County Sheriff's Office

Jail

Ship To:

Weld County Sheriff's Office

2875 West 10th Street

Greeley, CO 80634

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99949 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/25/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1001 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0996 g/210L

☒ Standard Deviation (<0.003) .0011

☒ Relative Standard Deviation (<3%) .0105

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0380130001I

Performed By: 732516 Date: 4/25/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/25/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1402

MP1681

☒ Printer

Reviewed By: 467497

Date: 4/26/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/23/2013

Serial Number: 90-000381

AssetTag: 775-22172

Facility

Weld County Sheriff's Office

Southwest Services

MAC Address:

Ship To:

IP Address:

Weld County Sheriff's Office

Gateway:

1950 O Street, Attn Rick Feit

Subnet Mask:

Greeley, CO 80631

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99971 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/23/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0963 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0959 g/210L

☒ Standard Deviation (<0.003) .0007

☒ Relative Standard Deviation (<3%) .0078

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0381130001I

Performed By: 467497 Date: 4/23/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/23/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1655

MP1329

☒ Printer

Reviewed By: 732516

Date: 4/23/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/9/2013
Serial Number: 90-000382
AssetTag: 775-22173

Facility

Westminster Police Department
Ship To:
Westminster Police Department
9110 Yates Street
Westminster, CO 80030

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99999 Pass ☒
Calibration Adjustment Pass ☒
Maintenance Log Entry ☒

Performed By: 467497 Date: 4/9/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1005 g/210L
☒ Linearity Test ☒ Interferant Detected
☒ Stability Test Average: 0.1007 g/210L
☒ Standard Deviation (<0.003) .0007
☒ Relative Standard Deviation (<3%) .0074
☒ EBAT Breath Test ☒ Final EBAT Test
☒ Improper Sample Test ☒ Maintenance Log Entry
☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com
☒ Card Reader ☒ Signature Pad
☒ IP Settings ☒ Modem
☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0382130001I

Performed By: 467497 Date: 4/9/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)
☒ Keyboard
☒ Signature Pad with USB Adapter
☒ Simulator Communication Cable
☒ Printer Communication Cable
☒ Trainer Access Card
☒ Power Supply Cables (3)
☒ Ethernet Cable
☒ Standard Simulator Solution

Lot Number:

Performed By: 467497
Date: 4/9/2013

Documents

☒ Certificate
☒ Exception Message Guide
☒ Operator Guide
☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1416 MP1632

☒ Printer

Reviewed By: 079490
Date: 4/20/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/22/2013

Serial Number: 90-000383

AssetTag: 775-22174

Facility

Wheat Ridge Police Department

Ship To:

Wheat Ridge Police Department

7500 West 29th Avenue

Wheatridge, CO 80215-6713

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99947 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/22/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0987 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0988 g/210L

☒ Standard Deviation (<0.003) .0008

☒ Relative Standard Deviation (<3%) .0084

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0383130001I

Performed By: 732516 Date: 4/22/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/22/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1654

MP1538

☒ Printer

Reviewed By: 079490

Date: 4/23/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/12/2013

Serial Number: 90-000384

AssetTag: 775-22175

Facility

Windsor Police Department

Ship To:

Windsor Police Department

200 N 11th Street

Windsor, CO 80550

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99867 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/12/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1022 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1007 g/210L

☒ Standard Deviation (<0.003) .0007

☒ Relative Standard Deviation (<3%) .0073

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0684130001I

Performed By: 467497 Date: 4/12/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/12/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1350

MP1518

☒ Printer

Reviewed By: 079490

Date: 4/16/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/24/2013

Serial Number: 90-000385

AssetTag: 775-22176

Facility

Woodland Park Police Department

Ship To:

Woodland Park Police Department

911 Tamarac Parkway

Woodland Park, CO 80863

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99894 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/24/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0970 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0965 g/210L

☒ Standard Deviation (<0.003) .0008

☒ Relative Standard Deviation (<3%) .0079

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0385130001I

Performed By: 732516 Date: 4/24/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/24/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1328

MP1315

☒ Printer

Reviewed By: 079490

Date: 4/25/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/17/2013

Serial Number: 90-000386

AssetTag: 775-22177

Facility

Yuma Police Department

Ship To:

Yuma Police Department

216 East 3rd Street

Yuma, CO 80759

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99882 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/17/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0991 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0992 g/210L

☒ Standard Deviation (<0.003) .0008

☒ Relative Standard Deviation (<3%) .0082

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0386130001I

Performed By: 732516 Date: 4/17/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/17/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1435

MP1299

☒ Printer

Reviewed By: 079490

Date: 4/17/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/16/2013

Serial Number: 90-000387

AssetTag: 775-22178

Facility

Hinsdale County Sheriff's Office

Ship To:

Hinsdale County Sheriff's Office

311 North Henson Street

Lake City, CO 81235

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99889 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 079490 Date: 4/16/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1009 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1009 g/210L

☒ Standard Deviation (<0.003) .0006

☒ Relative Standard Deviation (<3%) .0064

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0387130001I

Performed By: 079490 Date: 4/17/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 079490

Date: 4/17/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1590

MP1594

☒ Printer

Reviewed By: 467497

Date: 4/17/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/23/2013

Serial Number: 90-000388

AssetTag: 775-22179

Facility

San Juan County Sheriff's Office

Ship To:

San Juan County Sheriff's Office

1557 Greene Street

Silverton, CO 81433

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99987 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 079490 Date: 4/23/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0983 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0978 g/210L

☒ Standard Deviation (<0.003) .0007

☒ Relative Standard Deviation (<3%) .0073

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0388130001I

Performed By: 079490 Date: 4/23/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 079490

Date: 4/23/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1324

MP1331

☒ Printer

Reviewed By: 467497

Date: 4/23/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/23/2013

Serial Number: 90-000389

AssetTag: 775-22180

Facility

Yuma County Sheriff's Office

Ship To:

Yuma County Sheriff's Office

310 Ash Street

Wray, CO 80758

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99948 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/23/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0995 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0994 g/210L

☒ Standard Deviation (<0.003) .0008

☒ Relative Standard Deviation (<3%) .0082

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0389130001I

Performed By: 732516 Date: 4/23/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/23/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1436

MP1626

☒ Printer

Reviewed By: 467497

Date: 4/23/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/12/2013

Serial Number: 90-000412

AssetTag: 775-22181

Facility

Rio Blanco County County Sheriff's Office

Ship To:

Rio Blanco County County Sheriff's Office

555 Main Street

Meeker, CO 81641

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99825 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/12/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0985 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1003 g/210L

☒ Standard Deviation (<0.003) 0.0006

☒ Relative Standard Deviation (<3%) 0.0064

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0412130002I

Performed By: 467497 Date: 4/12/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/12/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1588

MP1354

☒ Printer

Reviewed By: 079490

Date: 4/16/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/25/2013

Serial Number: 90-000413

AssetTag: 775-22182

Facility

Sedgwick County Sheriff's Office

Ship To:

Sedgwick County Sheriff's Office

315 Ceder Street

Julesburg, CO 80737

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99968 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/25/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1003 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.1005 g/210L

☒ Standard Deviation (<0.003) .0008

☒ Relative Standard Deviation (<3%) .0082

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0413130001I

Performed By: 732516 Date: 4/25/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/25/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1650

MP1546

☒ Printer

Reviewed By: 079490

Date: 4/26/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/22/2013

Serial Number: 90-000414

AssetTag: 775-22183

Facility

Saguache County Sheriff's Office

Ship To:

Saguache County Sheriff's Office

530 5th Street

Saguache, CO 81149

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.999935 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/22/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0982 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0984 g/210L

☒ Standard Deviation (<0.003) .0005

☒ Relative Standard Deviation (<3%) .0051

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0414130001I

Performed By: 732516 Date: 4/22/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/22/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1471

MP1685

☒ Printer

Reviewed By: 079490

Date: 4/23/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/23/2013

Serial Number: 90-000415

AssetTag: 775-22184

Facility

Custer County Sheriff's Office

Ship To:

Custer County Sheriff's Office

702 Rosita Street

Westcliffe, CO 81252

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99956 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/23/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0979 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0978 g/210L

☒ Standard Deviation (<0.003) .0007

☒ Relative Standard Deviation (<3%) .0071

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0415130001I

Performed By: 732516 Date: 4/23/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/23/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1550

MP1367

☒ Printer

Reviewed By: 467497

Date: 4/23/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/16/2013

Serial Number: 90-000416

AssetTag: 775-22185

Facility

Montrose County Sheriff's Office

Ship To:

Montrose County Sheriff's Office

420 Horton

Olathe, CO 81425

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99991 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 079490 Date: 4/16/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0994 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0989 g/210L

☒ Standard Deviation (<0.003) .0008

☒ Relative Standard Deviation (<3%) .0080

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0416130001I

Performed By: 079490 Date: 4/17/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 079490

Date: 4/17/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1482

MP1447

☒ Printer

Reviewed By: 467497

Date: 4/20/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/5/2013
Serial Number: 90-000417
AssetTag: 775-22186

Facility

Aurora Police Department
Ship To:
Aurora Police Department
1290 South Potomac
Aurora, CO 80012

2 Arapahoe House

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99834 Pass ☒
Calibration Adjustment Pass ☒
Maintenance Log Entry ☒

Performed By: 467497 Date: 4/5/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1011 g/210L
☒ Linearity Test ☒ Interferant Detected
☒ Stability Test Average: 0.1006 g/210L
☒ Standard Deviation (<0.003) .0007
☒ Relative Standard Deviation (<3%) .0068
☒ EBAT Breath Test ☒ Final EBAT Test
☒ Improper Sample Test ☒ Maintenance Log Entry
☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com
☒ Card Reader ☒ Signature Pad
☒ IP Settings ☒ Modem
☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0417130001I

Performed By: 467497 Date: 4/5/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)
☒ Keyboard
☒ Signature Pad with USB Adapter
☒ Simulator Communication Cable
☒ Printer Communication Cable
☒ Trainer Access Card
☒ Power Supply Cables (3)
☒ Ethernet Cable
☒ Standard Simulator Solution

Lot Number:

Performed By: 467497
Date: 4/5/2013

Documents

☒ Certificate
☒ Exception Message Guide
☒ Operator Guide
☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1613 MP1440

☒ Printer

Reviewed By: 079490
Date: 4/20/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/26/2013

Serial Number: 90-000418

AssetTag: 775-22187

Facility

Aurora Police Department

3 Arapahoe House

Ship To:

Aurora Police Department

1290 South Potomac

Aurora, CO 80012

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99898 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/26/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0989 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0987 g/210L

☒ Standard Deviation (<0.003) .0009

☒ Relative Standard Deviation (<3%) .0088

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0418130001I

Performed By: 732516 Date: 4/26/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/26/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1369

MP1339

☒ Printer

Reviewed By: 079490

Date: 4/26/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/10/2013

Serial Number: 90-000419

AssetTag: 775-22188

Facility

Colorado Springs Police Department

Falcon

Ship To:

Colorado Springs Police Department

7850 Goddard Street

Colorado Springs, CO 80920

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99920 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/10/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1010 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0998 g/210L

☒ Standard Deviation (<0.003) .0007

☒ Relative Standard Deviation (<3%) .0070

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0419130001I

Performed By: 467497 Date: 4/11/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/11/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1666

MP1480

☒ Printer

Reviewed By: 079490

Date: 4/20/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/5/2013
Serial Number: 90-000420
AssetTag: 775-22189

Facility

Colorado Springs Police Department
Ship To:
Colorado Springs Police Department
955 West Moreno Avenue
Colorado Springs, CO 80905

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99985 Pass ☒
Calibration Adjustment Pass ☒
Maintenance Log Entry ☒

Performed By: 467497 Date: 4/5/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1005 g/210L
☒ Linearity Test ☒ Interferant Detected
☒ Stability Test Average: 0.0999 g/210L
☒ Standard Deviation (<0.003) .0007
☒ Relative Standard Deviation (<3%) .00072
☒ EBAT Breath Test ☒ Final EBAT Test
☒ Improper Sample Test ☒ Maintenance Log Entry
☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com
☒ Card Reader ☒ Signature Pad
☒ IP Settings ☒ Modem
☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0420130001I

Performed By: 467497 Date: 4/5/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)
☒ Keyboard
☒ Signature Pad with USB Adapter
☒ Simulator Communication Cable
☒ Printer Communication Cable
☒ Trainer Access Card
☒ Power Supply Cables (3)
☒ Ethernet Cable
☒ Standard Simulator Solution

Lot Number:

Performed By: 467497
Date: 4/5/2013

Documents

☒ Certificate
☒ Exception Message Guide
☒ Operator Guide
☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1659 MP1484

☒ Printer

Reviewed By: 079490
Date: 4/20/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 5/13/2013

Serial Number: 90-000422

AssetTag: 775-22191

Facility

Denver Police Department

Jail 1

Ship To:

Denver Police Department

3381 Park Avenue West

Denver, CO 80216

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) .99931 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 5/13/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0977 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0989 g/210L

☒ Standard Deviation (<0.003) 0.0006

☒ Relative Standard Deviation (<3%) 0.0061

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0422130001I

Performed By: 732516 Date: 5/13/2013

I-9000 Equipment Inventory

☐ Simulator Power Supplies (2)

☐ Keyboard

☐ Signature Pad with USB Adapter

☐ Simulator Communication Cable

☐ Printer Communication Cable

☐ Trainer Access Card

☐ Power Supply Cables (3)

☐ Ethernet Cable

☐ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 5/13/2013

Documents

☒ Certificate

☐ Exception Message Guide

☐ Operator Guide

☐ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

☐ Printer

Reviewed By: 467497

Date: 5/20/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/17/2013

Serial Number: 90-000423

AssetTag: 775-22192

Facility

Denver Police Department

Jail 2

Ship To:

Denver Police Department

3381 Park Avenue West

Denver, CO 80216

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99888 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/17/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0991 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0996 g/210L

☒ Standard Deviation (<0.003) .0007

☒ Relative Standard Deviation (<3%) .0069

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0423130001I

Performed By: 467497 Date: 4/17/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/17/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1404

MP1340

☒ Printer

Reviewed By: 079490

Date: 4/22/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/5/2013
Serial Number: 90-000424
AssetTag: 775-22193

Facility

Federal Heights Police Department
Ship To:
Federal Heights Police Department
2380 West 90th Avenue
Federal Heights, CO 80221

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99931 Pass ☒
Calibration Adjustment Pass ☒
Maintenance Log Entry ☒

Performed By: 467497 Date: 4/5/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.1009 g/210L
☒ Linearity Test ☒ Interferant Detected
☒ Stability Test Average: 0.1000 g/210L
☒ Standard Deviation (<0.003) .0007
☒ Relative Standard Deviation (<3%) .0073
☒ EBAT Breath Test ☒ Final EBAT Test
☒ Improper Sample Test ☒ Maintenance Log Entry
☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com
☒ Card Reader ☒ Signature Pad
☒ IP Settings ☒ Modem
☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0424130001I

Performed By: 467497 Date: 4/5/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)
☒ Keyboard
☒ Signature Pad with USB Adapter
☒ Simulator Communication Cable
☒ Printer Communication Cable
☒ Trainer Access Card
☒ Power Supply Cables (3)
☒ Ethernet Cable
☒ Standard Simulator Solution

Lot Number:

Performed By: 467497
Date: 4/5/2013

Documents

☒ Certificate
☒ Exception Message Guide
☒ Operator Guide
☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1651 MP1415

☒ Printer

Reviewed By: 079490
Date: 4/16/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/23/2013

Serial Number: 90-000426

AssetTag: 775-22195

Facility

Grand Junction Police Department
Ship To:
Grand Junction Police Department
555 Ute Avenue
Grand Junction, CO 81501

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99911 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 4/23/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0982 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0979 g/210L

☒ Standard Deviation (<0.003) .0006

☒ Relative Standard Deviation (<3%) .0065

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0426130001I

Performed By: 732516 Date: 4/23/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 732516

Date: 4/23/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1385

MP1636

☒ Printer

Reviewed By: 079490

Date: 4/23/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 5/13/2013

Serial Number: 90-000429

AssetTag: 775-22198

Facility

Calhan Police Department
Ship To:
Calhan Police Department
556 Colorado Avenue
Calhan, CO 80808

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99907 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 732516 Date: 5/13/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.0988 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0987 g/210L

☒ Standard Deviation (<0.003) 0.0009

☒ Relative Standard Deviation (<3%) 0.0088

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0429130001I

Performed By: 732516 Date: 5/13/2013

I-9000 Equipment Inventory

- ☐ Simulator Power Supplies (2)
- ☐ Keyboard
- ☐ Signature Pad with USB Adapter
- ☐ Simulator Communication Cable
- ☐ Printer Communication Cable
- ☐ Trainer Access Card
- ☐ Power Supply Cables (3)
- ☐ Ethernet Cable
- ☐ Standard Simulator Solution

Lot Number:

Performed By: 732516
Date: 5/13/2013

Documents

- ☒ Certificate
- ☐ Exception Message Guide
- ☐ Operator Guide
- ☐ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

- ☐ Printer

Reviewed By: 467497
Date: 5/20/2013

I9000 Certification Worksheet

Certification Type: Instrument Validation

Open Date: 4/17/2013

Serial Number: 90-000431

AssetTag: 775-22200

Facility

Brighton Police Department

Ship To:

Brighton Police Department

3401 East Bromley Lane

Brighton, CO 80601

MAC Address:

IP Address:

Gateway:

Subnet Mask:

Analog Phone:

Calibration Adjust

Flow Sensor (>0.998) 0.99875 Pass ☒

Calibration Adjustment Pass ☒

Maintenance Log Entry ☒

Performed By: 467497 Date: 4/17/2013

Calibration Verification

☒ Solution Change (0.095-0.105) Average: 0.99875 g/210L

☒ Linearity Test ☒ Interferant Detected

☒ Stability Test Average: 0.0994 g/210L

☒ Standard Deviation (<0.003) .0008

☒ Relative Standard Deviation (<3%) .0083

☒ EBAT Breath Test ☒ Final EBAT Test

☒ Improper Sample Test ☒ Maintenance Log Entry

☒ Invalid Sample Test ☒ I-9000 Performance Report

Function Checks

☒ Keyboard ☒ Printer Com

☒ Card Reader ☒ Signature Pad

☒ IP Settings ☒ Modem

☒ Field Settings ☒ Simulator Com

I-9000 Certification Certificate #

0431130001I

Performed By: 467497 Date: 4/17/2013

I-9000 Equipment Inventory

☒ Simulator Power Supplies (2)

☒ Keyboard

☒ Signature Pad with USB Adapter

☒ Simulator Communication Cable

☒ Printer Communication Cable

☒ Trainer Access Card

☒ Power Supply Cables (3)

☒ Ethernet Cable

☒ Standard Simulator Solution

Lot Number:

Performed By: 467497

Date: 4/17/2013

Documents

☒ Certificate

☒ Exception Message Guide

☒ Operator Guide

☒ Guth 12V500 Manual

Peripheral Equipment

Guth 12V500 Simulators (2)

MP1393

MP1376

☒ Printer

Reviewed By: 079490

Date: 4/17/2013